

CC-3427



'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob.: +91-9891121066, 9868309579 • E-mail: advance.cal@gmail.com

Calibration Certificate

CERTIFICATE NO.	: 230821A/N-104		Page	: 1 of 1
CALIBRATED ON.	: 21 / 08 / 2023		Ref. Std.	: CG-11
CALIBRATION DUE.	ON DUE. : 20 / 08 / 2024		C.P. No.	: ACS/CP/07
SRF No.& Dt.OF RECPT : 230821A/N Dt. 21.		.08.2023	Phy. Inspection : O.K.	
ULR No. & ISSUE DATE : CC342723000000		01125F Dt. 23.08.202	23 Cal. At	: Site / Lab
CALIBRATED FOR	: M/s SRL, Deptt. o J.N. Medical (f Microbiology. College, AMU Aligarh	1 -202002	
ITEM CALIBRATED	: Digital Temperatu	re Indicator of Refrig		
Make :			Range : (-20 to 50) °C	
	L. Count : 0.1			
	Location : 1 st	Floor	Deptt. : Microbiology	
ENVIRONMENT COM	NDITION: Temp. { 30	°C <u>+</u> 10°C} R.H. {	30% to 80%}	
STANDARD INSTRU	MENTS USED :			5V 10V
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Universal Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2022/0846	10/11/2023
RESULTS :	ET Discipline			4
* U. C. INSTT.		STANDARD	ERROR	
READS (°C)		READING (°C)	(°C)	
MUT CALL		3		
2.0		4	0.1	
6.1		6	0.1	
8.12 A S		8	0.1	
		rage of readings)		

REMARKS

The Expanded uncertainty of measurement at 95% confidence level ± 1.35°C at coverage Factor k=1.96

*U.C. Under Calibration.

Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above

The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above

This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services

4. The Calibration Certificate is not to be used for any legal purpose and snall not be produced in the count of last.

5. Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use



