

Parashar Micro Measurement Pvt. Ltd.



B-59, SECTOR-64, NOIDA (U.P.) 201 307
 PH.: 0120-4252360-65, Email : pmmpl.india@gmail.com
 Website : www.pmmpl.in, www.parashar.info



Calibration Certificate

ULR - CC279524000025096F		Page 1 of 1				
Certificate Number : PMM/270424/11-1		Date of Request	25.04.2024			
Calibrated For	M/s. AHM & Dufferin Female Hospital Pathology Lab. Kanpur.	Calibrated on	27.04.2024			
		Suggested Due Date	26.04.2025			
		Cert. Issue Date	29.04.2024			
Description of Equipments						
Nomenclature	Centrifuge	Accuracy/Class	Not Mentioned			
Range	Upto 6000 rpm	Condition of UUC	Physically Ok			
Resolution / Least Count	10 rpm	Serial No.	ZHKN-43328			
Make & Model No.	Remi / R-8C Plus	Party ID Mark No.	-----			
Material / Type	-----	Location	-----			
Master Equipment / Standard Used						
Sr.No.	Nomenclature	Make & Model	Serial No. / I/D Mark	Calibrated From	Certificate No.	Due Date of Calibration
1	Dig. Tachometer	Metravi	210813911 PMM/TM/09	PMMPL, Noida	PMM/1529	01.12.2024
Standard used for calibration are traceable to Accredited lab, for Standard ISO/IEC : 17025 - 2017 or National Standards Through unbroken chain of calibration						
Calibration Performed at		Site	Environmental	Temperature = 23 ± 5° C		
Calibration Procedure No.		CP/M/MV/S/03	Condition(s)	Humidity = 30 to 75 % RH		
Reference Standard / Guideline		-----				
Mechanical Calibration						
CALIBRATION RESULTS						
Sr.No.	Nominal Value in UUC (in rpm)	Actual Observed Value in Std. (in rpm)	Error (in rpm)	± Uncertainty in Measurement (in rpm)		
1	1000	1000	0	13		
2	2000	2001	-1	13		
3	3000	3002	-2	13		
4	4000	4003	-3	13		
5	5800	5805	-5	13		

UUC - Unit Under Calibration

Std. - Standard Instrument

Uncertainty of Measurement (at approx 95% Confidence Level with Coverage factor $k = 2$)

Aakash

Calibrated By : Aakash Kumar
 Designation : Calibration Engineer



Jai Prakash

Approved By : Jai Prakash
 Designation : Tech. Manager / Auth. Sign.

Conditions : 1. This certificate refers only to the particular item submitted for calibration.
 2. The calibration result reported in this certificate are valid at the time of and under the stated conditions of measurement.
 3. This particular certificate can not be reproduced except in full, without prior permission of chief executive officer of the lab.