



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2648

Certificate No	AACPL / 05971F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240827.3	DT - 27 / 08 / 2024		ULR No.	CC2646240000 05971F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited			Calibration Date	27.08.2024		
Address	Regional Hospital Dhalpur Kullu H.P			Recommended Date for Next Calib.	26.08.2025		
				Certificate Issue Dt.	31.08.2024		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No	Party ID / Location
1	Micro Pipette	5 to 50 µl	0.5 µl	Dragon Lab YE222AX0053027	KDLHPKLCC01 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	55 %RH	23.5 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Error (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	20.0	20.00158	-0.00158	0.25
2	30.0	30.00193	-0.00193	0.25
3	50.0	50.00215	-0.00215	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Anand Chaurasia
(CEO)