



# AACPL CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2648

Certificate No	AACPL / 05973F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240827.3	DT - 27 / 08 / 2024		ULR No.	CC2646240000 05973F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited			Calibration Date	27.08.2024		
Address	Regional Hospital Dhalpur Kullu H.P			Recommended Date for Next Calib.	26.08.2025		
				Certificate Issue Dt.	31.08.2024		
				UUC Condition	Good		

### Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Sl.No	Party ID / Location
1	Micro Pipette	100 to 1000 µl	5 µl	Dragon Lab YE221AW0154680	KDLHPKLCC03 Clinical Chemistry

### Standard Equipments Used ( Traceable to National Standard )

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6	Volume Measurement	55 %RH	23.5 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

### Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Error (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	100	99.99846	0.00154	0.25
2	500	499.99786	0.00214	0.25
3	1000	999.99723	0.00277	0.25

#### Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by  
  
Anand Chaurasia  
(CEO)