



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 05975F	Field	Thermal	Group	Temp. & RH	Page No.	1 of 1
SRF No & Date	240827.3	DT - 27 / 08 / 2024		ULR No.	CC2646240000 05975F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited			Calibration Date	27.08.2024		
Address	Regional Hospital Dhalpur Kullu H.P			Recommended Date for Next Callb.	26.08.2025		
				Certificate Issue Dt.	31.08.2024		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Dig. Thermo Hygrometer	-10 to 50 °C 10 to 99 % RH	0.1 °C / 1 % RH	HTC -----	KDLHPKLC01 Clinical Pathology

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI.No	Calibrated At	Cal Certificate No	Cal. Valid Upto
1	Dig.Temp. & Humidity Ind. With sensor	Polltech 5019/20275542	Belz	30077505	07/02/2025
2	Humidity Chamber	Indecon / -----	-----	-----	-----

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
Comparison Method	Temp. & RH Measurement	52 %RH	25.4 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 03, Section No : CPTH - 07

Calibration Points :- Temp. 15, 25 & 40 °C Relative Humidity 30, 60 & 90 % RH

Calibration Results (Temperature at 50 % RH)

Sl. No.	Master Value on Std. (In °C)	UUC Value (In °C)	Deviation (In °C)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (°C)
1	15.06	14.9	-0.16	0.46
2	25.13	24.7	-0.43	0.46
3	40.17	39.6	-0.57	0.46

Calibration Results (Humidity at 25 °C)

Sl. No.	Master Value on Std. (In % RH)	UUC Value (In % RH)	Deviation (In % RH)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (%RH)
1	30.05	30	-0.05	1.64
2	60.09	60	-0.09	1.64
3	90.12	90	-0.12	1.64

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

 Anand Chaurasia
 (CEO)