



AACALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 05998F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	240828.1	DT - 28 / 08 / 2024		ULR No.	CC2646240000 05998F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited			Calibration Date	28.08.2024		
Address	Regional Hospital Dhalpur Kullu H.P.			Recommended Date for Next Calib.	27.08.2025		
				Certificate Issue Dt.	31.08.2024		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Sl.No.	Party ID / Location
1	Centrifuge M/c.	880 RPM	NA	Matrix BHCC-271/22-23	KDLHPKLHM03 Hematology

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Start Temp.	End Temp.	Cal. Performed At
IS : 12508 - 2018	RPM Measurement	56 %RH	24.4 °C	24.9 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

Calibration Result

Sl	Calib. Gauge Value	Master Gauge Value	Error	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in RPM)
	UUC (in RPM)	(in RPM)	(in RPM)	
1	880	880.5	-0.5	3.90

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Anand Chaurasia
(CEO)