



CALIBRATION CERTIFICATE



CC - 2646

Certificate No	AACPL / 05992F	Field	Thermal	Group	Temp.	Page No.	1 of 1
SRF No & Date	240828.1	DT - 28 / 08 / 2024		ULR No.	CC2646240000 05992F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited			Calibration Date	28.08.2024		
Address	Regional Hospital Dhalpur Kullu H.P			Recommended Date for Next Calib.	27.08.2025		
				Certificate Issue Dt.	31.08.2024		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No	Party ID / Location
1	Refrigerator	2 to 8 °C	1 °C	Noorcol HP-KDL62656/OE/288	KDLHPKLHM02 Hematology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Digital Multimeter	Gwinstek EL 160577	AACPL	AACPL/05131F	09/07/2025
2	RTD (4 Wire) Pt-100	Tempens/ 115	AACPL	AACPL/07063F	27/09/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
DKD-R5-7	Temp. Measurement	57 %RH	27.4 °C	Site

Calibration Procedure

Comparison With Standard Sensors as per Calibration Procedure Manual AACPL / CP / 02, Section No : CPTH - 09

Calibration Points :- 2, 5 & 7 °C

Calibration Results

SI No.	Master Value	UUC Value	Error In °C	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (°C)
	In °C	In °C		
1	2.05	2	-0.05	0.65
2	5.07	5	-0.07	0.65
3	7.12	7	-0.12	0.65

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration
- Temperature Scale: International Temperature Scale Of 1990 (ITS-90)

----- End of Certificate -----



 Calibrated by

 Approved By

 Anand Chaurasia
 (CEO)