



CHRISTIANMEDICALCOLLEGE
VELLORE, TAMILNADU-632004.INDIA
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMCEXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT-JULY 2024

LABNO: 8672
LABNAME: **JEEYAR INTEGRATIVE MEDICAL
SERVICES (JIMS) HOSPITAL**

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 17/07/2024

Date of Result Published: 08/08/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method/ Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-PODII	ERBA ANALYZER CHEM-5 Plus V2	90	161.02	8.28	13.34	167 mg/dL	0.45	2.81
2	UREA II	Urease UV/GLDHII	ERBA ANALYZER CHEM-5 Plus V2	72	57.21	15.19	8.69	57 mg/dL	-0.02	2.05
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	ERBA ANALYZER CHEM-5 Plus V2	72	2.99	10.95	0.33	3.4 mg/dL	1.24	0.08
4	T.BILIRUBIN II	Diazonium Salt(Colorimetric)/ Jendrassik II	ERBA ANALYZER CHEM-5 Plus V2	84	3.9	10.56	0.41	4.1 mg/dL	0.49	0.09
5	T-PROTEIN II	Biuret- Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	83	5.24	9.03	0.47	5.1 g/dL	-0.3	0.1
6	ALBUMIN II	BCG-Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	81	3.17	7.64	0.24	3.5 g/dL	1.38	0.05
7	CHOLESTROL II	CHOD-PAPII	ERBA ANALYZER CHEM-5 Plus V2	86	117.55	10.7	12.57	105 mg/dL	-1	2.71
8	TRIGLYCERIDE II	GPO-PAP/ Enzymatic Colorimetric/ End Point II	ERBA ANALYZER CHEM-5 Plus V2	88	110.53	15.15	16.74	128 mg/dL	1.04	3.57

u*-method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable(Action Signal)

LABADDRESS

JEEYAR INTEGRATIVE
MEDICAL SERVICES
(JIMS) HOSPITAL, JIVA
Campus, Muchintal Road,
Hyderabad

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in
tactNo.:0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - JUNE 2024

LAB NO: 8672

LAB NAME: **JEEYAR INTEGRATIVE MEDICAL
SERVICES (JIMS) HOSPITAL**

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 14/06/2024

Date of Result Published: 27/06/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	ERBA ANALYZER CHEM-5 Plus V2	94	268.31	6.56	17.6	232 mg/dL	-2.06	3.63
2	UREA II	Urease UV / GLDH II	ERBA ANALYZER CHEM-5 Plus V2	75	103.96	11.3	11.75	112 mg/dL	0.68	2.71
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	ERBA ANALYZER CHEM-5 Plus V2	77	4.23	10.68	0.45	4.4 mg/dL	0.38	0.1
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	ERBA ANALYZER CHEM-5 Plus V2	83	2.22	13.33	0.3	2.2 mg/dL	-0.07	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	79	5.04	11.66	0.59	4.2 g/dL	-1.42	0.13
6	ALBUMIN II	BCG - Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	83	3.08	9.79	0.3	2.8 g/dL	-0.93	0.07
7	CHOLESTROL II	CHOD-PAP II	ERBA ANALYZER CHEM-5 Plus V2	81	104.6	8.79	9.2	114 mg/dL	1.02	2.04
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	ERBA ANALYZER CHEM-5 Plus V2	86	176.5	9.13	16.12	162 mg/dL	-0.9	3.48

u* - method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

LAB ADDRESS

JEEYAR INTEGRATIVE
MEDICAL SERVICES
(JIMS) HOSPITAL, JIVA
Campus, Muchintal Road,
Hyderabad

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - MAY 2024

LAB NO: 8672

LAB NAME: **JEEYAR INTEGRATIVE MEDICAL
SERVICES (JIMS) HOSPITAL**

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 16/05/2024

Date of Result Published: 30/05/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	ERBA ANALYZER CHEM-5 Plus V2	88	126.02	8.56	10.79	109 mg/dL	-1.58	2.3
2	UREA II	Urease UV / GLDH II	ERBA ANALYZER CHEM-5 Plus V2	67	76.6	10.49	8.04	82 mg/dL	0.67	1.96
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	ERBA ANALYZER CHEM-5 Plus V2	73	1.82	12.42	0.23	1.92 mg/dL	0.43	0.05
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	ERBA ANALYZER CHEM-5 Plus V2	78	3.42	11.89	0.41	2.59 mg/dL	-2.02	0.09
5	T-PROTEIN II	Biuret - Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	76	5.08	12.66	0.64	4.7 g/dL	-0.59	0.15
6	ALBUMIN II	BCG - Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	80	3.13	11.61	0.36	2.9 g/dL	-0.64	0.08
7	CHOLESTROL II	CHOD-PAP II	ERBA ANALYZER CHEM-5 Plus V2	84	107.49	10.67	11.47	96 mg/dL	-1	2.5
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	ERBA ANALYZER CHEM-5 Plus V2	84	88.42	13.26	11.73	106 mg/dL	1.5	2.56

u* - method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

LAB ADDRESS

JEEYAR INTEGRATIVE
MEDICAL SERVICES
(JIMS) HOSPITAL, JIVA
Campus, Muchintal Road,
Hyderabad

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.



CHRISTIANMEDICALCOLLEGE

VELLORE, TAMILNADU-632004.INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT-APRIL 2024

LABNO: 8672

LABNAME: **JEEYAR INTEGRATIVE MEDICAL SERVICES (JIMS) HOSPITAL**

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 17/04/2024

PT Item: Lyophilized human serum based

Date of Result Published: 08/05/2024

S. No.	Analyte	Method/ Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	ERBA ANALYZER CHEM-5 Plus V2	82	206.16	8.19	16.88	201 mg/dL	-0.31	3.73
2	UREA II	Urease UV/GLDH II	ERBA ANALYZER CHEM-5 Plus V2	69	37.46	12.4	4.64	32 mg/dL	-1.18	1.12
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	ERBA ANALYZER CHEM-5 Plus V2	72	3.73	10.1	0.38	4.7 mg/dL	2.55	0.09
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric)/ Jendrassik II	ERBA ANALYZER CHEM-5 Plus V2	79	5.97	9.08	0.54	5.9 mg/dL	-0.13	0.12
5	T-PROTEIN II	Biuret-Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	76	5.17	13.01	0.67	6.2 g/dL	1.54	0.15
6	ALBUMIN II	BCG-Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	78	3.16	8.79	0.28	2.1 g/dL	-3.79	0.06
7	CHOLESTROL II	CHOD-PAP II	ERBA ANALYZER CHEM-5 Plus V2	85	111.12	9.99	11.1	122 mg/dL	0.98	2.41
8	TRIGLYCERIDE II	GPO-PAP/ Enzymatic Colorimetric/End Point II	ERBA ANALYZER CHEM-5 Plus V2	84	342.2	11.04	37.76	341 mg/dL	-0.03	8.24

u*-method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

LAB ADDRESS

JEEYAR INTEGRATIVE
MEDICAL SERVICES (JIMS)
HOSPITAL, JIVA Campus,
Muchintal Road,
Hyderabad

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.



CHRISTIANMEDICALCOLLEGE

VELLORE, TAMILNADU-632004.INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - MARCH 2024

LABNO: 8672

LABNAME: **JEEYAR INTEGRATIVE MEDICAL SERVICES (JIMS) HOSPITAL**

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 20/03/2024

Date of Result Published: 06/04/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method/ Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	ERBA ANALYZER CHEM-5 Plus V2	84	402.7	8.73	35.14	307 mg/dL	-2.72	7.67
2	UREA II	UreaseUV/GLDH II	ERBA ANALYZER CHEM-5 Plus V2	68	66.28	8.84	5.86	79 mg/dL	2.17	1.42
3	CREATININE II	Jaffes Kinetic- Alkaline Picrate II	ERBA ANALYZER CHEM-5 Plus V2	66	7.12	10.15	0.72	8.1 mg/dL	1.36	0.18
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric)/ Jendrassik II	ERBA ANALYZER CHEM-5 Plus V2	81	2.51	13.64	0.34	1.87 mg/dL	-1.88	0.08
5	T-PROTEIN II	Biuret-Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	69	3.99	10.37	0.41	4.6 g/dL	1.49	0.1
6	ALBUMIN II	BCG-Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	70	2.54	10.22	0.26	2.7 g/dL	0.62	0.06
7	CHOLESTROL II	CHOD-PAPII	ERBA ANALYZER CHEM-5 Plus V2	75	85.71	9.04	7.75	89 mg/dL	0.42	1.79
8	TRIGLYCERIDE II	GPO-PAP/ Enzymatic Colorimetric/End Point II	ERBA ANALYZER CHEM-5 Plus V2	75	74.44	13.49	10.04	66 mg/dL	-0.84	2.32

u*-method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

LAB ADDRESS

JEEYAR INTEGRATIVE
MEDICAL SERVICES (JIMS)
HOSPITAL, JIVA Campus,
Muchintal Road,
Hyderabad

Coordinator Contact Details:

E-
Mail: clinqc@cmcvellore.ac.in
Cont
act No.: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - FEBRUARY 2024

LAB NO: 8672

LAB NAME: **JEEYAR INTEGRATIVE MEDICAL
SERVICES (JIMS) HOSPITAL**

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 09/02/2024

Date of Result Published: 05/03/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	ERBA ANALYZER CHEM-5 Plus V2	81	94.56	7.74	7.32	69 mg/dL	-3.49	1.63
2	UREA II	Urease UV / GLDH II	ERBA ANALYZER CHEM-5 Plus V2	62	26.26	11.67	3.07	48 mg/dL	7.08	0.78
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	ERBA ANALYZER CHEM-5 Plus V2	72	2.33	13.42	0.31	2.1 mg/dL	-0.74	0.07
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	ERBA ANALYZER CHEM-5 Plus V2	70	1.57	16.67	0.26	3.1 mg/dL	5.88	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	73	5.72	11.48	0.66	6.2 g/dL	0.73	0.15
6	ALBUMIN II	BCG - Colorimetric II	ERBA ANALYZER CHEM-5 Plus V2	73	3.56	9.89	0.35	2.1 g/dL	-4.17	0.08
7	CHOLESTROL II	CHOD-PAP II	ERBA ANALYZER CHEM-5 Plus V2	84	126.41	11.15	14.09	87 mg/dL	-2.8	3.08
8	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	ERBA ANALYZER CHEM-5 Plus V2	79	161.36	8.87	14.31	121 mg/dL	-2.82	3.22

u* - method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

LAB ADDRESS

JEEYAR INTEGRATIVE
MEDICAL SERVICES
(JIMS) HOSPITAL, JIVA
Campus, Muchintal Road,
Hyderabad

Coordinator Contact Details:

E-Mail: clinqc@cmcvellore.ac.in
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential.
CMC EQAS does not sub contract any components.
***** End of Report *****

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.