



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,MOGA (1054)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 02

Report Date : 10/07/2024

| Analyte | Instrument | Result Value | Standard Unit | Z-Score |
|---|------------|--------------|---------------|---------|
| No Outlier | | | | |
| Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation. | | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

| Analyte | Standard Unit | Result Value | Mean | Z-Score | RMZ |
|--------------------|---------------|--------------|--------|---------|-------|
| @ Glucose | mg/dl | 78.80 | 71.18 | 0.10 | 0.73 |
| @ Urea | mg/dl | 25.00 | 26.57 | -0.13 | -0.34 |
| @ Creatinine | mg/dl | 1.01 | 0.99 | 0.16 | -0.33 |
| @ Bilirubin Total | mg/dl | 0.20 | 0.35 | -0.63 | 0.97 |
| @ Bilirubin Direct | mg/dl | 0.19 | 0.14 | 1.90 | 1.17 |
| @ Total Protein | g/dL | 7.65 | 7.58 | 0.46 | -0.63 |
| @ Albumin | g/dL | 4.47 | 4.19 | 1.04 | 1.01 |
| @ Uric Acid | mg/dl | 4.41 | 4.32 | -0.23 | -0.36 |
| @ Calcium | mg/dl | 8.52 | 8.55 | -0.24 | 0.10 |
| @ Triglycerides | mg/dl | 148.00 | 135.91 | 0.57 | -0.04 |
| @ HDL | mg/dl | 29.00 | 29.24 | -0.50 | 0.58 |
| @ LDL | mg/dl | 92.37 | 89.65 | -0.23 | -0.89 |
| @ ALT | U/L | 7.30 | 5.92 | 0.96 | 0.11 |
| @ AST | U/L | 16.94 | 13.94 | 0.71 | 0.54 |
| @ ALP | U/L | 51.80 | 45.36 | -0.24 | -0.56 |

Instrument : Diasys Sys 200


| Analyte | Standard Unit | Result Value | Mean | Z-Score | RMZ |
|---------------|---------------|--------------|--------|---------|-------|
| @ Cholesterol | mg/dl | 164.00 | 169.63 | -0.34 | -0.29 |

Instrument : ACCULYTE-3P

| Analyte | Standard Unit | Result Value | Mean | Z-Score | RMZ |
|-------------|---------------|--------------|--------|---------|-------|
| @ Chloride | mmol/L | 104.12 | 105.34 | 0.22 | -0.14 |
| @ Sodium | mmol/L | 128.13 | 133.36 | -0.47 | -0.75 |
| @ Potassium | mmol/L | 4.98 | 5.13 | -0.39 | -0.44 |

- Legend @ : Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score ≥3 - Unacceptable
 # : Not Evaluated
 ⌚ : Delayed Result Entry
 * : Not considered for evaluation.

| | |
|--------------------------|----------|
| Total Parameters | 19 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 19 |
| Outlier Parameters (X) | 0 |
| EQAS Score Biochemistry | 100.00 % |



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070