



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,MOGA (1054)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	40.80	40.67	0.84	-0.34
@ Creatinine	mg/dl	0.77	0.85	-0.94	-0.33
@ Bilirubin Total	mg/dl	0.29	0.45	-0.68	0.97
@ Bilirubin Direct	mg/dl	0.28	0.22	0.50	1.17
@ Total Protein	g/dL	6.75	6.81	0.22	-0.63
@ Albumin	g/dL	4.27	4.20	0.63	1.01
@ Uric Acid	mg/dl	4.25	4.36	-0.53	-0.36
@ Calcium	mg/dl	8.26	8.82	-1.08	0.10
@ Triglycerides	mg/dl	127.00	137.36	-1.11	-0.04
@ HDL	mg/dl	25.00	31.25	-1.48	0.58
@ LDL	mg/dl	97.66	99.01	-0.38	-0.89
@ ALT	U/L	19.10	21.41	-0.08	0.11
@ AST	U/L	39.07	41.16	0.24	0.54
@ ALP	U/L	75.88	70.28	-0.37	-0.56

Instrument : Diasys Sys 200

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Cholesterol	mg/dl	172.00	166.55	-0.04	-0.29

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	111.61	114.30	-0.12	-0.14
@ Sodium	mmol/L	162.98	162.40	0.06	-0.75
@ Potassium	mmol/L	5.97	6.31	-0.45	-0.44

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥ 3 - Unacceptable

: Not Evaluated

⌚ : Delayed Result Entry

* : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	0
Evaluated Parameters	18
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS

Unit No. 409-416.

Commercial Building - 1A

Kohinoor Mall, Kirol Road, Kurla (W),

Mumbai - 400070