



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD,MOGA (1054)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

| Analyte | Instrument | Result Value | Standard Unit |
|--|------------|--------------|---------------|
| X Widal test - Typhi H antigen (TH) | Manual | 1:80 | - |
| Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry | | | |

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Manual


| Analyte | Standard Unit | Result Value | Accepted Value |
|--|---------------|------------------|------------------|
| X Widal test - Typhi H antigen (TH) | - | 1:80 | No Agglutination |
| @ RPR (Syphilis Serology) | - | Non-reactive | Non-reactive |
| @ Widal test - Typhi O antigen (TO) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi A-H antigen (AH) | - | No Agglutination | No Agglutination |
| @ Widal test - ParaTyphi B-H antigen (BH) | - | No Agglutination | No Agglutination |

Instrument : J Mitra

| Analyte | Standard Unit | Result Value | Accepted Value |
|----------|---------------|--------------|----------------|
| @ HBsAg | IU/mL | Non-reactive | Non-reactive |
| @ HCV | IU/mL | Non-reactive | Non-reactive |
| @ HIV Ab | S/CO | Non-reactive | Non-reactive |

Legend @ : Acceptable
 @ : Acceptable
 X : Unacceptable
 # : Not Evaluated
 Ⓞ : Delayed Result Entry

| | |
|---------------------------------|---------|
| Total Parameters | 8 |
| Not Evaluated Parameters | 0 |
| Evaluated Parameters | 8 |
| Outlier Parameters (X) | 1 |
| EQAS Score Immunoassay Extended | 87.50 % |



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070