



# A A CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 05952F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240824.3	DT - 24 / 08 / 2024		ULR No.	CC2646240000 05952F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	24.08.2024		
Address	Inside Civil Hospital Kapurthala (P.B)			Recommended Date for Next Calib.	23.08.2025		
				Certificate Issue Dt.	27.08.2024		
				UUC Condition	Good		

### Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	20 to 200 µl	1 µl	Dragon Lab YE222AX0047966	KDPBKPMMMP03 Hematology

### Standard Equipments Used ( Traceable to National Standard )

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	56 %RH	23.7 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

### Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Error (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is $\pm$ (in µl)
1	20	20.10111	-0.10111	0.25
2	100	100.33569	-0.33569	0.25
3	200	200.78698	-0.78698	0.25

### Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

2  
Kai  
Tomar  
Calibrated by



Approved by  
  
Anand Chaurasia  
(CEO)