



# A A CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 07063F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	230926.5 DT - 26 / 09 / 2023			ULR No.	CC2646230000 07063F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Ltd.			Calibration Date	26.09.2023		
Address	Mathura Daas Civil Hospital (DH), Main Bazar, Moga (PB)			Recommended Date for Next Calib.	25.09.2024		
				Certificate Issue Dt.	29.09.2023		
				UUC Condition	Good		

### Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI. No	Party ID / Location
1	Centrifuge M/c.	0 to 6000 RPM	10 RPM	Remi -----	KDDHPBCPCF07 Clinical Pathology

### Standard Equipments Used ( Traceable to National Standard )

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 06363F	31/08/2024

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
IS : 12508 - 2018	RPM Measurement	(50 ± 20)%RH	(25 ± 10) °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

### Calibration Result

SI	Calib. Gauge Value	Master Gauge Value	Deviation	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is $\pm$ (in RPM)
	UUC ( in RPM )	( in RPM )	( in RPM )	
1	100	100.1	-0.1	5.88
2	500	500.2	-0.2	5.88
3	1000	1000.2	-0.2	5.88
4	3000	3000.4	-0.4	6.15
5	6000	6000.6	-0.6	6.17

### Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

*Saty Jany*  
Calibrated by



Approved by  
*Anand Chaurasia*  
Anand Chaurasia  
(CEO)

LAB :- Plot No. C-18, Sector-13, First Floor, "Vasundhara, Haat" Vasundhara, Ghaziabad - 201012

Ph. : 0120-4135529, Mob. : 9212172564, 9212172597

F-mail : aacplinfo2005@gmail.com, aacplinfo@rediffmail.com