

**CALIBRATION REPORT**

ADL/DO/MS/C/00207/24-25

Customer: Shifa Hospital, Adirampattinam
Serial No.: 211930113714

Model : Mispa Viva

DOI: 29-03-2019

Date of Calibration: 29/04/24

INSTRUMENT CONTROL

Visual Control: PASSED

Zero Drift after 1 Hour of zero setting (WL=340nm) 0.001

PHOTOMETRIC LINEARITY

Solution	WL (nm)	100%	75%	50%	25%
Potassium dichromate	340nm	100.2	74.6	51.8	25.2
P-Nitrophenol	405nm	100.4	74.5	50.3	25.8
Cobaltous Ammonium Sulphate	505nm	100.8	76.9	50.1	24.7

CHEMISTRY CONTROL

	TEST	REAGENT	CONTROL SERUM	THEORETICAL VALUE	READ VALUE	RANGE
END POINT	GLUCOSE	AGAPPE	Lot No: 32403648 Expiry: 28/2/26	96mg/dl	92	82-110
KINETIC	SGOT	AGAPPE	Lot No: 32403648 Expiry: 28/2/26	35.4 IU/l	37	26.9-43.9
FIXED TIME	CREAT	AGAPPE	Lot No: 32403648 Expiry: 28/2/26	0.94 mg/dl	1.01	0.59-1.29

Validity: 6 months from the date of calibration.

For Agappe Diagnostics Ltd.,

Authorized Signatory

Authorized Signatory

Name: H. Sriram

Designation: Asst. Manager - Customer service

AGAPPE DIAGNOSTICS LTD. ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413**CORPORATE OFFICE / REAGENT PLANT**Agappe Hills, Pattanamattom (PO), Dist. Ernakulam, Kerala - 681562, India.
Tel: +91 484 286 7000 | Email: agappe@agappe.in**EQUIPMENT PLANT**K/585-CB, Block no. 32, KINFRA Small Industrial Park,
Nettali, Cochin, Kerala, India - 686721. Tel: +91 484 276 7477.**KOLKATA OFFICE**406, Merlin Matrix, Plot No-10, Block-09K, Sector V,
Salt Lake City, Kolkata - 700 091.
Tel: +91 33 4003 0451 | Email: kolkataoffice@agappe.in**MUMBAI (REGISTERED OFFICE)**401 & 402, 4th Floor, Jaasigh Business Centre, 119,
Sahar Road, Parsiwada, Andheri (East), Mumbai - 400 099, India.
Tel: +91 22 4300 8000 | Email: mumbaioffice@agappe.in**DELHI OFFICE**DSM 540, 5th Floor, DLF Tower, Viharji Marg,
New Delhi - 110 015, India.
Tel: +91 11 4508 8416 | Email: delhioffice@agappe.in**BANGALORE OFFICE**5-6, Bth Floor, Red Cross Bhavan, No-26 Race Course Road,
Bangalore - 560 001. Tel: +91 80 2228 8288
Email: bangaloreoffice@agappe.in