



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

## DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



### MONTHLY SUMMARY REPORT - AUGUST 2024

LAB NO: 10193

LAB NAME: **QRIS HEALTH**

PT Item: Lyophilized human whole blood based

CONSTITUENT GROUP

**HbA1c**

Date of Result Entered: 10/08/2024

Date of Result Published: 02/09/2024

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	243	5.29	9.04	0.48	4.9 %	-0.81	0.06

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (Action Signal)

#### LAB ADDRESS


QRIS HEALTH,  
#24/25, SECTOR-3, ROHINI,  
Delhi

#### Coordinator Contact Details:

E-Mail: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential.  
CMC EQAS does not sub contract any components.  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*  
Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

## DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

### MONTHLY SUMMARY REPORT - JULY 2024

LAB NO: 10193  
LAB NAME: **QRIS HEALTH**

PT Item: Lyophilized human whole blood based

CONSTITUENT GROUP

**HbA1c**

Date of Result Entered: 09/07/2024

Date of Result Published: 08/08/2024

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	229	6.14	9.35	0.57	6.8 %	1.16	0.08

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (Action Signal)

#### LAB ADDRESS


QRIS HEALTH,  
F24/25, SECTOR-3, ROHINI,  
Delhi

#### Coordinator Contact Details:

E-Mail: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential.  
CMC EQAS does not sub contract any components.  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*  
Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

## DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

### MONTHLY SUMMARY REPORT - JUNE 2024

LAB NO: 10193

LAB NAME: **CURIS DIAGNOSTICS PVT LTD**

PT Item: Lyophilized human whole blood based

CONSTITUENT GROUP

**HbA1c**

Date of Result Entered: 03/06/2024

Date of Result Published: 27/06/2024

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	223	5.21	8.98	0.47	5.2 %	-0.02	0.06

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (Action Signal)

#### LAB ADDRESS

CURIS DIAGNOSTICS PVT LTD,  
F24/25, SECTOR-3, ROHINI,  
Delhi

#### Coordinator Contact Details:

E-Mail: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential.  
CMC EQAS does not sub contract any components.  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*

Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

## DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

### MONTHLY SUMMARY REPORT - MAY 2024

LAB NO: 10193

LAB NAME: **CURIS DIAGNOSTICS PVT LTD**

PT Item: Lyophilized human whole blood based

CONSTITUENT GROUP

**HbA1c**

Date of Result Entered: 02/05/2024

Date of Result Published: 30/05/2024

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	209	5.77	9.65	0.56	6.6 %	1.49	0.08

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (Action Signal)

#### LAB ADDRESS

CURIS DIAGNOSTICS PVT LTD,  
F24/25, SECTOR-3, ROHINI,  
Delhi

#### Coordinator Contact Details:

E-Mail: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential.  
CMC EQAS does not sub contract any components.  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*

Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

## DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

### MONTHLY SUMMARY REPORT - APRIL 2024

LAB NO: 10193

LAB NAME: **CURIS DIAGNOSTICS PVT LTD**

PT Item: Lyophilized human whole blood based

CONSTITUENT GROUP

**HbA1c**

Date of Result Entered: 03/04/2024

Date of Result Published: 08/05/2024

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	202	5.19	8.47	0.44	5.2 %	0.02	0.06

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (Action Signal)

#### LAB ADDRESS

CURIS DIAGNOSTICS PVT LTD,  
F24/25, SECTOR-3, ROHINI,  
Delhi


#### Coordinator Contact Details:

E-Mail: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential.  
CMC EQAS does not sub contract any components.  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*

Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

## DEPARTMENT OF CLINICAL BIOCHEMISTRY

### CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

### MONTHLY SUMMARY REPORT - MARCH 2024

LAB NO: 10193

LAB NAME: **CURIS DIAGNOSTICS PVT LTD**

PT Item: Lyophilized human whole blood based

CONSTITUENT GROUP

**HbA1c**

Date of Result Entered: 03/03/2024

Date of Result Published: 06/04/2024

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Any Analyser	193	6.93	8.07	0.56	7.7 %	1.38	0.08

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (Action Signal)

#### LAB ADDRESS

CURIS DIAGNOSTICS PVT LTD,  
F24/25, SECTOR-3, ROHINI,  
Delhi

#### Coordinator Contact Details:

E-Mail: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential.  
CMC EQAS does not sub contract any components.  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*

Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.