

Corrective Action and Preventive Action (CAPA)

MAE - EQAS outlier.

Sr. No.	Date	Nature of Problem	Root Cause, Corrective and Preventive Actions Taken	Date of closure	Signature
1	20/3/24	WBC was out - 4.97.	Got ILC done results were comparable	20/3/24	<i>Devendra</i>
2	21/6/24	RBC & HB were 5.96 & 4.53 could be random error	Got ILC done results were ok. To be monitored in next cycle.	21/6/24	<i>Devendra</i>