



**HORIBA India Private Ltd.**

246, Okhla Industrial

Estate, Phase - III 110

020 New Delhi, India

Tel: +91 (11) 4646 5000 / 4669 5001

Fax: +91 (11) 4669 5010 / 4646 5020

HIN/MED/2024-2025/1476

**22<sup>th</sup> August 2024.**

**CALIBRATION CERTIFICATE**

This is to certify that the Hematology Analyzer YH550 bearing serial number: 109YAXH03504 installed Curis Diagnostics pvt ltd was Calibrated on **22<sup>th</sup> August 2024.**

Calibrator: ABX MINOCAL

Lot No.: CX498

Expiry Date: 05<sup>th</sup> October 2024.

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

**Next calibration cycle is due on 21<sup>th</sup> August 2025.**



**Shrish Dixit**

(Head- Products & Marketing)

For **Horiba India Pvt. Ltd.**