



A A CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06276F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	240904.4	DT - 04 / 09 / 2024		ULR No.	CC2646240000 06276F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited			Calibration Date	04.09.2024		
Address	Inside Civil Hospital Ferozepur, Dist.-Ferozepur-152002			Recommended Date for Next Calib.	03.09.2025		
				Certificate Issue Dt.	06.09.2024		
				UUC Condition	Good		

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Centrifuge M/c.	0 to 6000 RPM	10 RPM	Remi -----	KDPBFZCCCN02 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 05973F	26/08/2025

Reference Standard	Calibration Purpose	Humidity	Start Temp.	End Temp.	Cal. Performed At
IS : 12508 - 2018	RPM Measurement	54 %RH	25.2 °C	25.4 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

Calibration Result

Sl	Calib. Gauge Value UUC (in RPM)	Master Gauge Value (in RPM)	Error (in RPM)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in RPM)	
1	100	99.9	0.1		5.88
2	500	499.8	0.2		5.88
3	1500	1499.6	0.4		5.88
4	2500	2499.4	0.6		6.15
5	3500	3499.1	0.9		6.15
6	5000	4998.7	1.3		6.15

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----


Calibrated by



Approved by

Anand Chaurasia
(CEO)