



# AACALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06278F	Field	Mech.	Group	Speed	Page No.	1 of 1
SRF No & Date	240904.4 DT - 04 / 09 / 2024			ULR No.	CC2646240000 06278F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited Inside Civil Hospital Ferozepur, Dist.-Ferozepur-152002			Calibration Date	04.09.2024		
Address				Recommended Date for Next Calib.	03.09.2025		
				Certificate Issue Dt.	06.09.2024		
				UUC Condition	Good		

### Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Sl.No.	Party ID / Location
1	Centrifuge M/c.	0 to 5000 RPM	100 RPM	Weswox Optik 76960	KDPBFZCPCN03 Clinical Pathology

### Standard Equipments Used ( Traceable to National Standard )

Sl	Instrument Name	Make/Sl No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Dig. Tachometer	Line Seiki / -----	AACPL	AACPL / 05973F	26/08/2025

Reference Standard	Calibration Purpose	Humidity	Start Temp.	End Temp.	Cal. Performed At
IS : 12508 - 2018	RPM Measurement	54 %RH	25.2 °C	25.4 °C	Site

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01B, Section No : CP - 08

### Calibration Result

Sl	Calib. Gauge Value UUC ( in RPM )	Master Gauge Value ( in RPM )	Error ( in RPM )	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in RPM)	
1	100	100.4	-0.4	57.75	
2	500	500.7	-0.7	57.75	
3	1500	1501.2	-1.2	57.75	
4	2500	2501.9	-1.9	57.77	
5	3500	3502.8	-2.8	57.77	
6	5000	5003.6	-3.6	57.77	

**Note**

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by  
  
Anand Chaurasia  
(CEO)