



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06266F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240904.2 DT - 04 / 09 / 2024	ULR No.		CC2646240000 06266F			
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited Inside Civil Hospital Ferozepur, Dist.-Ferozepur-152002			Calibration Date		04.09.2024	
Address				Recommended Date for Next Calib.		03.09.2025	
				Certificate Issue Dt.		06.09.2024	
				UUC Condition		Good	

Calibration Instrument Detail

Sl	Instrument Name	Range	Least Count	Make / Sl.No.	Party ID / Location
1	Micro Pipette	100 to 1000 µl	5 µl	Dragon Lab YE221AW0154510	KDPBFZCCMP01 Clinical Chemistry

Standard Equipments Used (Traceable to National Standard)

Sl	Instrument Name	Make/Sl No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	54 %RH	24.2 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Error (in µl)	Uncertainty at approx 95% confidence Level and coverage factor k = 2 is ± (in µl)
1	100	100.02351	-0.02351	0.25
2	500	500.04767	-0.04767	0.25
3	1000	1000.07139	-0.07139	0.25

Note

- 1 The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- 2 This certificate cannot be reproduced except in full without our prior permission in writing
- 3 This certificate refers only to the particular item(s) submitted for calibration
- 4 UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Anand Chaurasia
(CEO)