



AACPL CALIBRATION PVT. LTD.

CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06268F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240904.2	DT - 04 / 09 / 2024		ULR No.	CC2646240000 06268F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited			Calibration Date	04.09.2024		
Address	Inside Civil Hospital Ferozepur, Dist.-Ferozepur-152002			Recommended Date for Next Calib.	03.09.2025		
				Certificate Issue Dt.	06.09.2024		
				UUC Condition	Good		

Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	20 to 200 μ l	1 μ l	Dragon Lab YE222AX0047986	KDPBFZHMP03 Haematology

Standard Equipments Used (Traceable to National Standard)

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	55 %RH	25.1 $^{\circ}$ C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

Calibration Result

S. No	Nominal Value on UUC (in μ l)	Average Volume at 27 $^{\circ}$ C (in μ l)	Error (in μ l)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is \pm (in μ l)
1	20	20.00573	-0.00573	0.25
2	100	100.02461	-0.02461	0.25
3	200	200.05196	-0.05196	0.25

Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by

Anand Chaurasia
(CEO)