



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, FEROZPUR (1045)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro


Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	42.00	40.67	1.07	2.65
@ Creatinine	mg/dl	0.70	0.85	-1.34	0.78
@ Bilirubin Total	mg/dl	0.49	0.45	0.62	0.01
@ Bilirubin Direct	mg/dl	0.23	0.22	0.00	-0.13
@ Total Protein	g/dL	7.03	6.81	0.74	0.54
@ Albumin	g/dL	4.10	4.20	0.00	0.92
@ Uric Acid	mg/dl	4.14	4.36	-0.79	0.55
@ Calcium	mg/dl	8.76	8.82	-0.04	0.18
@ Cholesterol	mg/dl	167.00	170.36	-0.43	0.60
@ Triglycerides	mg/dl	138.00	137.36	-0.25	-0.44
@ HDL	mg/dl	38.00	31.25	0.27	0.71
@ LDL	mg/dl	90.38	99.01	-1.14	-0.68
@ ALT	U/L	18.60	21.41	-0.18	1.34
@ AST	U/L	38.98	41.16	0.23	2.86
@ ALP	U/L	62.05	70.28	-1.31	1.47

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	115.30	114.30	0.41	1.02
@ Sodium	mmol/L	162.39	162.40	0.00	2.17
@ Potassium	mmol/L	6.55	6.31	0.50	2.72

Legend @ : Acceptable
! : 2.0 < z score < 3.0 - Warning
X : z score ≥3 - Unacceptable
: Not Evaluated
⊙ : Delayed Result Entry
* : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	0
Evaluated Parameters	18
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070