



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, FEROZPUR (1045)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
X Specific Gravity	Manual	1.015	-
X Blood (urine)	Manual	Trace	-

Legend @ : Acceptable
X : Unacceptable
: Not Evaluated
⌚ : Delayed Result Entry

Problem Classification: _____

Corrective Action: _____


Reviewed by: _____

Dated: _____

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
X Specific Gravity	-	1.015	1.020,1.025,1.030
X Blood (urine)	-	Trace	+1,+2,+3
@ pH	-	5	5,5.5,6
@ Glucose (urine)	-	2+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,+3
@ Ketone (urine)	-	Negative	+1,Neg,Tr
@ Bilirubin (urine)	-	Negative	+1,Neg
@ Urobilinogen (urine)	-	Normal	+1,Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable	Total Parameters	9
@ : Acceptable	Not Evaluated Parameters	0
X : Unacceptable	Evaluated Parameters	9
# : Not Evaluated	Outlier Parameters (X)	2
⌚ : Delayed Result Entry	EQAS Score Clinical Path - Urine Routine Chemistry	77.78 %



Dr Puneet Kumar Nigam
PT coordinator & Technical Manager, MHL EQAS
Unit No. 409-416.
Commercial Building - 1A
Kohinoor Mall, Kirol Road, Kurla (W),
Mumbai - 400070



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS LTD, FEROZPUR (1045)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry			

Problem Classification: _____


Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value										
@ Crystal	-	Triple phosphate crystal	Triple phosphate crystal										
Legend @ : Acceptable @ : Acceptable X : Unacceptable # : Not Evaluated Ⓞ : Delayed Result Entry		<table border="1"> <tr> <td>Total Parameters</td> <td>1</td> </tr> <tr> <td>Not Evaluated Parameters</td> <td>0</td> </tr> <tr> <td>Evaluated Parameters</td> <td>1</td> </tr> <tr> <td>Outlier Parameters (X)</td> <td>0</td> </tr> <tr> <td>EQAS Score Clinical Path - Urine Sediment Analysis Image</td> <td>100.00 %</td> </tr> </table>		Total Parameters	1	Not Evaluated Parameters	0	Evaluated Parameters	1	Outlier Parameters (X)	0	EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %
Total Parameters	1												
Not Evaluated Parameters	0												
Evaluated Parameters	1												
Outlier Parameters (X)	0												
EQAS Score Clinical Path - Urine Sediment Analysis Image	100.00 %												



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kiroil Road, Kurla (W),
 Mumbai - 400070



Krsnaa Diagnostics Ltd
DH, Housing Board Colony,
Ferozpur PB -152001

INTER LABORATORY COMPARISON RESULTS

Sr. No	Date	Analyte Name	Barcode ID	Lab Result	Unit	ILC lab Result	Unit	Acceptability Criteria	Observed CV %	Acceptability range	Result Acceptable Yes / No	Done by	Reviewed by
1.	23/7/24	Specific Gravity	39766	1.010	-	1.015	-	10%	-0.49%	-	Yes	[Signature]	[Signature]

Qualitative: Clinically comparable results shall be Concordance.
Clinically not comparable results shall be Disconcordance.

Quantitative: Acceptability: 10 %. (As per NABL 112)


[Signature]
Sign of Doctor

Issued by : Vinod Lonkar <u>[Signature]</u>		Issue Date : 01.03.2024	
Approved by :			
Doc No. GEN-01	Version No : 2	Amend No. --	Amend Date .




Medical Laboratory Report

Patient Name : Mr VEER SINGH
Age and Gender : 27 Years / Male
Category : OPD - PB FEROZEPUR
Referring Doctor : CIVIL HOSPITAL FEROZEPUR CITY
Sample Processed at : PB FEROZEPUR

Patient UID No : PFP240700039766
PRN No : PFP240700039766
Registered On : 23 Jul 2024 12:47
Sample UID No.  39766

CLINICAL PATHOLOGY

Test Done	Observed Value	Units	Biological Reference Interval
ROUTINE EXAMINATION OF URINE			
PHYSICAL EXAMINATION (Sample : Urine)			
Quantity	20	ml.	
Colour	Pale Yellow		Pale Yellow
Deposit	Absent		Absent
Appearance	Clear		Clear
CHEMICAL EXAMINATION			
pH	6.5		4.60 - 8.00
Bromocresol Green			
Specific Gravity	1.010		1.010 - 1.030
Bromothymol Blue			
Protein	Absent		Absent
Protein Error Principle			
Sugar	Absent		Absent
GOD - POD			
Urine Ketone	Absent		Absent
Legals test			
Blood	Absent		Absent
Dihydro Peroxidase Tetramethyl Benzidine Method			
Bile salts	Absent		Absent
Hays Sulphur Powder			
Bile Pigments	Absent		Absent
Azo Diazo Reaction			
MICROSCOPIC EXAMINATION			
Red Blood Cells	Absent		0-2
Pus Cells	1-2	hpf	0-5
Epithelial Cells	0-1	hpf	0-5
Crystals	Absent		Absent
Casts	Absent		
Other	Absent		



END OF REPORT

Sample Collected On : 23.07.2024 12:57
Results Authenticated : 23.07.2024 13:36

Sample Accepted On : 23.07.2024 13:03
Results Reported : 23.07.2024 13:45

Printed On : 11.09.2024 14:59
E19387





Laboratory Diagnostics Services on Public Private Partnership Mode
Punjab Health Systems Corporations, Govt. of Punjab
with Krsnaa Diagnostics (Mohali) Pvt. Ltd.



Medical Laboratory Report

Patient Name : Mr VEER SINGH
Age and Gender : 27 Years / Male
Category : OPD - PB FEROPUR
Referring Doctor : CIVIL HOSPITAL FEROPUR CITY
Sample Processed at : KDMPLCHPBJALANDHAR144001

Patient UID No : PFP240700039765
PRN No : PFP240700039765
Registered On : 23 Jul 2024 12:45
Sample UID No. 39765

CLINICAL PATHOLOGY

Test Done	Observed Value	Units	Biological Reference Interval
ROUTINE EXAMINATION OF URINE			
PHYSICAL EXAMINATION (Sample : Urine)			
Quantity	15	ml.	
Colour	Pale Yellow		Pale Yellow
Deposit	Absent		Absent
Appearance	Clear		Clear
CHEMICAL EXAMINATION			
pH	6.5		4.60 - 8.00
Bromocresal Green			
Specific Gravity	1.015		1.010 - 1.030
Bromothymol Blue			
Protein	Absent		Absent
tetrabromphenol blue buffer nonreactive ingredients			
Sugar	Absent		Absent
glucose oxidase peroxidase potassium iodide buffer nonreactive ingredients			
Urine Ketone	Absent		Absent
sodium nitroprusside buffer nonreactive ingredients			
Blood	Absent		Absent
diisopropylbenzene dihydro peroxide tetramethylbenzidine buffer nonreactive ingredients			
Bile salts	Absent		Absent
2,4-dichloroaniline diazoninum salt buffer nonreactive ingredients			
Bile Pigments	Absent		Absent
2,4-dichloroaniline diazoninum salt buffer nonreactive ingredients			
MICROSCOPIC EXAMINATION			
Red Blood Cells	Absent	/ hpf	0-2 / hpf
Pus Cells	2-3	/ hpf	0-5 / hpf
Epithelial Cells	0-1		0-2 / hpf
Crystals	Absent		Absent

Devinder Singh
DR. DEVINDER SINGH VIRDI
(MD PATHO)



--- END OF REPORT ---

Sample Collected On : 23.07.2024 12:55
Results Authenticated : 24.07.2024 17:06

Sample Accepted On : 24.07.2024 17:05
Results Reported : 24.07.2024 17:31

E18141
Printed On : 11.09.2024 14:58


Krsnaa Diagnostics (Mohali) Pvt. Ltd.
All Pathology Services
Helpline No.: 020 8814 6814 | 020 4695 4695 | 96233 96233 | Email : info@krsnadiagnostics.com





Medical Laboratory Report

Patient Name : Mr VEER SINGH
Age and Gender : 27 Years / Male
Category : OPD - PB FEROZEPUR
Referring Doctor : CIVIL HOSPITAL FEROZEPUR CITY
Sample Processed at : KDMPLCHPBJALANDHAR144001

Patient UID No : PFP240700039765
PRN No : PFP240700039765
Registered On : 23 Jul 2024 12:45
Sample UID No.  39765

CLINICAL PATHOLOGY

Test Done	Observed Value	Units	Biological Reference Interval
-----------	----------------	-------	-------------------------------

ROUTINE EXAMINATION OF URINE

MICROSCOPIC EXAMINATION

Casts	Absent
Other	Absent

METHOD - REFLECTANCE PHOTOMETRY / MANUAL
MULTISTICK



Devinder Singh

DR.DEVINDER SINGH VIRDI
(MD PATHO)

--- END OF REPORT ---

Sample Collected On : 23.07.2024 12:55
Results Authenticated : 24.07.2024 17:06

Sample Accepted On : 24.07.2024 17:05
Results Reported : 24.07.2024 17:31

E18141
Printed On : 11.09.2024 14:58