



Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD, MANSA (1038)

Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score \leq 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score \geq 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

Problem Classification: _____

Corrective Action: _____

Reviewed by: _____

Dated: _____

Instrument : Diasys 200 Pro


Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Urea	mg/dl	34.80	40.67	-0.28	-0.16
@ Creatinine	mg/dl	0.97	0.85	0.21	0.22
@ Bilirubin Total	mg/dl	0.48	0.45	0.55	0.23
@ Bilirubin Direct	mg/dl	0.21	0.22	-0.20	0.33
@ Total Protein	g/dL	6.62	6.81	-0.02	0.25
@ Albumin	g/dL	4.10	4.20	0.00	0.75
@ Uric Acid	mg/dl	4.89	4.36	1.00	0.05
@ Calcium	mg/dl	8.77	8.82	-0.01	0.68
@ Cholesterol	mg/dl	160.00	170.36	-0.98	-0.26
@ Triglycerides	mg/dl	139.00	137.36	-0.17	0.40
@ HDL	mg/dl	22.00	31.25	-1.89	0.02
@ LDL	mg/dl	104.64	99.01	0.35	-0.04
@ ALT	U/L	20.40	21.41	0.18	0.70
@ AST	U/L	34.50	41.16	-0.27	0.52
@ ALP	U/L	54.41	70.28	-1.84	-0.92

Instrument : ACCULYTE-3P

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Chloride	mmol/L	115.61	114.30	0.46	0.48
@ Sodium	mmol/L	160.00	162.40	-0.22	-0.23
@ Potassium	mmol/L	6.45	6.31	0.34	-0.01

Legend @ : Acceptable
 ! : 2.0 < z score < 3.0 - Warning
 X : z score ≥3 - Unacceptable
 # : Not Evaluated
 ⊙ : Delayed Result Entry
 * : Not considered for evaluation.

Total Parameters	18
Not Evaluated Parameters	0
Evaluated Parameters	18
Outlier Parameters (X)	0
EQAS Score Biochemistry	100.00 %



Dr Puneet Kumar Nigam
 PT coordinator & Technical Manager, MHL EQAS
 Unit No. 409-416.
 Commercial Building - 1A
 Kohinoor Mall, Kirol Road, Kurla (W),
 Mumbai - 400070