



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD, MANSA (1038)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit	Z-Score
No Outlier				
Legend @ : z score $\leq$ 2.0 - Acceptable ! : 2.0 < z score < 3.0 - Warning X : z score $\geq$ 3 - Unacceptable # : Not Evaluated ⌚ : Delayed Result Entry * : Not considered for evaluation.				

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_

**Dated:** \_\_\_\_\_

Instrument : Horiba Yumizen H550

Analyte	Standard Unit	Result Value	Mean	Z-Score	RMZ
@ Erythrocyte (RBC) Count	mill/cu.mm	2.68	2.47	0.58	0.53
@ Total Leucocytes (WBC) count	/c.mm	2990.00	2649.14	0.59	0.35
@ Haemoglobin	g/dL	7.00	6.75	-0.19	-0.07
@ MCV (Mean Corpuscular Volume)	fL	107.90	106.32	-0.02	0.36
@ MCH (Mean Corpuscular Hb)	pg	26.30	27.25	-1.26	-0.61
@ MCHC (Mean Corpuscular Hb Concentration)	g/dL	24.40	25.58	-0.32	-0.66
@ Platelet Count	1000/ µL	141.00	158.56	-0.04	-0.33
@ Hematocrit	%	28.90	26.17	0.00	0.51
@ RDW CV%	%	19.10	18.60	0.28	0.82

Legend @ : Acceptable

! : 2.0 < z score < 3.0 - Warning


X : z score ≥3 - Unacceptable

# : Not Evaluated

⌚ : Delayed Result Entry

\* : Not considered for evaluation.

Total Parameters	9
Not Evaluated Parameters	0
Evaluated Parameters	9
Outlier Parameters (X)	0
EQAS Score Haematology - CBC	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD, MANSA (1038)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 03

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_


**Dated:** \_\_\_\_\_

Instrument : -

Analyte	Standard Unit	Result Value	Accepted Value
@ Diagnosis (Slide)	-	CLL/Chronic lymphoproliferative disorder	CLL/Chronic lymphoproliferative disorder

Legend @ : Acceptable  
 @ : Acceptable  
 X : Unacceptable  
 # : Not Evaluated  
 ⌚ : Delayed Result Entry

Total Parameters	1
Not Evaluated Parameters	0
Evaluated Parameters	1
Outlier Parameters (X)	0
EQAS Score Haematology- Peripheral Smear Diagnosis	100.00 %



**Dr Puneet Kumar Nigam**  
 PT coordinator & Technical Manager, MHL EQAS  
 Unit No. 409-416.  
 Commercial Building - 1A  
 Kohinoor Mall, Kiroi Road, Kurla (W),  
 Mumbai - 400070