



# Metropolis Healthcare Ltd., EQAS

KRSNAA DIAGNOSTICS,LTD, MANSA (1038)

## Outlier And Analyte Summary Report

Outlier Details For Cycle No 240104 and Sample No 01

Report Date : 10/07/2024

Analyte	Instrument	Result Value	Standard Unit
No Outlier			
Legend @ : Acceptable X : Unacceptable # : Not Evaluated ⌚ Delayed Result Entry			

**Problem Classification:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Corrective Action:** \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Reviewed by:** \_\_\_\_\_


**Dated:** \_\_\_\_\_

Instrument : Manual

Analyte	Standard Unit	Result Value	Accepted Value
@ pH	-	5	5,5.5,6
@ Specific Gravity	-	1.025	1.020,1.025,1.030
@ Glucose (urine)	-	3+	+2,+3,+4
@ Protein (Urine)	-	1+	+1,+2,+3
@ Ketone (urine)	-	Negative	+1,Neg,Tr
@ Blood (urine)	-	2+	+1,+2,+3
@ Bilirubin (urine)	-	Negative	+1,Neg
@ Urobilinogen (urine)	-	Normal	+1,Nor
@ Nitrite (urine)	-	Present	Pre

Legend @ : Acceptable	Total Parameters	9
@ : Acceptable	Not Evaluated Parameters	0
X : Unacceptable	Evaluated Parameters	9
# : Not Evaluated	Outlier Parameters (X)	0
⌚ : Delayed Result Entry	EQAS Score Clinical Path - Urine Routine Chemistry	100.00 %



**Dr Puneet Kumar Nigam**  
PT coordinator & Technical Manager, MHL EQAS  
Unit No. 409-416.  
Commercial Building - 1A  
Kohinoor Mall, Kirol Road, Kurla (W),  
Mumbai - 400070