



# AACPL CALIBRATION PVT. LTD.

## CALIBRATION CERTIFICATE

7.8F-01



CC - 2646

Certificate No	AACPL / 06307F	Field	Mech.	Group	Volume	Page No.	1 of 1
SRF No & Date	240905.3 DT - 05 / 09 / 2024			ULR No.	CC2646240000 06307F		
Calibration Method :	Comparison Method			Accuracy			
Company Name	M/s Krsnaa Diagnostics Limited			Calibration Date	05.09.2024		
Address	Inside Civil Hospital, Mansa (P.B)			Recommended Date for Next Calib.	04.09.2025		
				Certificate Issue Dt.	07.09.2024		
				UUC Condition	Good		

### Calibration Instrument Detail

SI	Instrument Name	Range	Least Count	Make / SI.No.	Party ID / Location
1	Micro Pipette	10 to 100 µl	1 µl	I-Pette -----	KDPBMADHSE11 Serology

### Standard Equipments Used ( Traceable to National Standard )

SI	Instrument Name	Make/SI No	Calibrated By	Cal Certificate No	Cal. Valid Upto
1	Weighing Balance	Radwag / 668683	Bluebox Techno	BBT/744/JAN/24	04/03/2025

Reference Standard	Calibration Purpose	Humidity	Temperature	Calibration Performed At
ISO - 8655 - 6 - 2022	Volume Measurement	55 %RH	24.3 °C	Laboratory

Calibration Procedure : As per Calibration Procedure Manual AACPL / CP / 01C, Section No : CP - 03

### Calibration Result

S. No	Nominal Value on UUC (in µl)	Average Volume at 27 °C (in µl)	Error (in µl)	Uncertainty at approx 95% confidence Level and coverage factor $k = 2$ is $\pm$ (in µl)
1	20	20.00981	-0.00981	0.25
2	50	50.02673	-0.02673	0.25
3	100	100.04152	-0.04152	0.25

#### Note

- The calibration results reported in this calibration certificate are valid at the time of & under stated condition
- This certificate cannot be reproduced except in full without our prior permission in writing
- This certificate refers only to the particular item(s) submitted for calibration
- UUC - Unit under calibration

----- End of Certificate -----

Calibrated by



Approved by  
  
Anand Chaurasia  
(CEO)