

CALIBRATION CERTIFICATE

Avantor Performance Materials India Pvt Limited

19th Floor, Building No. 5, Tower C DLF Cyber
City Phase - III, Gurgaon - 122001;
Haryana Phone: +91-124-465670;
Web: avantormaterials.com

IVD/CAL/C61-1/300524

MAY 30, 2024

NAME OF THE CUSTOMER Lakshmi Diagnostic Center, Panruti-607106.	MANUFACTURER AVANTOR PERFORMANCE MATERIALS INDIA PVT LTD
PRODUCT NAME BENESPHERA C61	SERIAL No: 602122023IE

This is to certify that the above mentioned equipment has been inspected and Calibrated as per the protocol on 30TH MAY, 2024. The Equipment is working satisfactorily. This Calibration is valid through 30th MAY, 2025. The Next Calibration is due on 29TH MAY, 2025.

S.No	TEST PARAMETERS	MEASURED/CALIBRATED	REMARKS
01	INPUT VOLTAGE	230V	PASSED
02	POWER CONSUMPTION	300VA	PASSED
03	INPUT FREQUENCY	50 HZ	PASSED
04	OPERATING TEMPERATURE	15 - 25 °C	PASSED

S.No	TEST PARAMETERS	STATUS TEST	REMARKS
01	LAMP CHECK	PASSED	WORKING
02	PUMP MODULE TEST	PASSED	WORKING
03	FILTER MODULE TEST	PASSED	WORKING
04	A/D INPUT TEST	PASSED	WORKING

For Avantor Performance Materials India Pvt Ltd

R.GOKULNATH
Area Service Manager.

Service Centre: #725 Ground Floor, Udyog Vihar Phase-V, Gurgaon - 122016, Haryana

IVD/CAL/H30_0524

May 30, 2024


<p>NAME OF THE CUSTOMER</p> <p>Lakshmi Diagnostic Center, Panruti - 607106.</p>	<p>MANUFACTURER</p> <p>MINDRAY</p>
<p>PRODUCT NAME</p> <p>Avantor H31 Plus</p>	<p>SERIAL No:</p> <p>UG-39001183</p>

This is to certify that the above mentioned equipment has been inspected and Calibrated as per the protocol on 30th MAY,2024. The Equipment is working satisfactorily. This Calibration is valid through 30th May,2025. The Next Calibration is due on 29th May,2025.

S.No	TEST PARAMETERS	MEASURED/CALIBRATED	REMARKS
01	INPUT VOLTAGE	230V	PASSED
02	POWER CONSUMPTION	300VA	PASSED
03	INPUT FREQUENCY	50 HZ	PASSED
04	OPERATING TEMPERATURE	15 - 25 *C	PASSED

S.No	TEST PARAMETERS	STATUS TEST	REMARKS
01 :	COUNTING TIME	PASSED	WORKING
02	SYRINGE MODULE	PASSED	WORKING
03	ROTATION MOTOR TEST	PASSED	WORKING
04	ELEVATION MOTOR TEST	PASSED	WORKING
05	VACCUM	204PSI	WORKING
07	VALVE CHECK	PASSED	WORKING
08	A/D INPUT CHECK	PASSED	WORKING

For Avantor Performance Materials India Limited,



R.GOKULNATH
Area Service Manager.