

CALIBRATION CERTIFICATE

This is to certify that the Hematology Analyzer **Yumizen H550** bearing serial number: **204YAXH03811** installed at **KRSNAA DIAGNOSTICS LTD, FEROZPUR, Punjab** has calibrated on 07 October 2023.

Calibrator : ABX MINOCAL
Lot No. : CX495
Expiry Date: 05th July 2024

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on **06 OCTOBER 2024.**

A circular stamp with the text "H. RIBA IN PVT. LTD." around the perimeter. Overlaid on the stamp is a handwritten signature in black ink.

Shrish Dixit

(Head- Products & Marketing)
For **Horiba India Pvt. Ltd.**