

CERTIFICATE OF CALIBRATION

SRF No.	F0131	Certificate No.	FCL/24/F0131-01	ULR No.	CC31032400000721F
Date of SRF	03-09-2024	Date of Cal.	03-09-2024	Next Cal Due	02-09-2025
Issue Date	05-09-2024	Cal Procedure No.	FCL-SOP-THE-01	Reference Std.	DKD-R 5-1
No of Pages	01	Page No.	01 of 01	Status on Receipt	Satisfactory
DISCIPLINE-GROUP : Thermal-Temperature				Cal Done At	Onsite

Customer Name & Full Address:

M/s. Nagamangala Government Hospital
Government Hospital
B M road , Nagamangala-571432
Mandya District

DUC DETAILS

Nomenclature	ILR Refrigerator	Model / Type	NA
Make	Godrej	Range	2 to 8 °C
Serial. No	220300027MR00247	Resolution	0.1 °C
ID No	ICTC/GHN/ILR-01	Accuracy	NA
Location	ICTC LAB		

STANDARDS USED FOR CALIBRATION AND TRACEABILITY DETAILS

Nomenclature	Make/Model	SI. No / ID No.	Traceable to	Validity
4 Wire RTD Sensor With DMM	Tempsens & Yokogawa / PT - 100X1, Simplex & CA71	2113016632 & 2267021/FCL-TH-02 & FCL-TH-02A	TMS/24/64-01	15-03-2025

Temperature Bath Calibrator used as Source

Environmental Condition :

Temperature : 25.4 °C

Humidity: 52 %RH

Calibration Results:

All Readings are in °C

SI No	Set Temp	STD Readings	DUC Readings	Observed Deviation	Computed Measurement Uncertainty (±)
1	5	5.351	5.0	-0.351	0.8

Remarks :

- The Measurement Uncertainty is estimated at a confidence level of 95.45% with a coverage factor k=2.
- This Calibration Certificate refers only to the Particular item submitted for calibration. Calibration points given as per customer request.
- NABL Accredited Calibration Lab as per ISO/IEC 17025:2017 with vide Certificate No:CC-3103
- The Usage of NABL symbol is as per NABL guidelines given in NABL 133. The Calibration Certificate relates only to the above DUC.
- Calibration Certificate Shall not be reproduced except in full, without written approval of the Flowcal.
- Standard maintained are traceable to National / International Standard through accredited laboratories.

Calibrated by



Sanju

(Calibration Engineer)



Authorized by



Vinay Kumar

(CEO)

*****End of Calibration Certificate*****

