



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 4266

Distribution No.: 163-K

Month/Year: April/2024

Instrument ID: Transasia

Model Name.: H560

Serial No.: K1104B2312034

Name & Contact No. of PT Co-ordinator: Dr. Manoranjan Mahapatra (Prof. & Head), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : info@ishtmaiimseqap.com

Date of issue & status of the report: 19-06-2024[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	4.44	4.38	8.82	8.55	0.033	0.32	0.06	0.1	0.006	-0.45
RBC x10 ⁶ /µl	1	4.57	4.56	9.13	9.1	0.010	0.12	0.01	0.04	0.003	-0.58
Hb g/dl	1	13	12.9	25.9	25.5	0.030	0.49	0.1	0.1	0.007	0.00
HCT%	1	37.5	37.5	75	79.2	0.170	-0.97	0	0.4	0.023	-1.08
MCV-fl	1	82.3	82.3	164.6	174	0.314	-1.15	0	0.3	0.022	-0.81
MCH-Pg	1	28.6	28.6	57.2	56.2	0.067	0.62	0	0.2	0.016	-0.90
MCHC-g/dl	1	34.8	34.8	69.6	64.35	0.161	1.25	0	0.3	0.019	-1.01
Plt. x10 ³ /µl	1	58	57	115	179.5	1.782	-1.32	1	5	0.297	-0.90
Retic %	2	19.7	13.8	33.5	16	0.260	2.46	5.9	0.5	0.036	9.11

P.S . Assesment

YOUR REPORT			CONSENSUS REPORT	
DLC%	3	Nrbcs=0 , Poly=8 L=2, E=2, Mono/Promono=1 , B1=68 P.M.=3, Mye=0, Meta=0, Other=0	Lymph: 77-88, Poly: 8-14.25, mono: 1-3, nRBC/Blast/Myelo/Meta/Eosino: 0-5	
RBC Morphology	3	Normocytic Normochromic Mostly	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, Mild: Microcytic	
Diagnosis	3		Chronic Lymphoproliferative Disorder/CLL	

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 163--K	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	303	303	79.21	89.44	5.61	5.28	15.18	5.28
RBC x10⁶/µl	1	303	303	89.44	90.76	5.28	3.96	5.28	5.28
Hb g/dl	1	303	303	90.1	91.42	5.61	3.96	4.29	4.62
HCT%	1	303	302	93.05	89.4	4.97	5.3	1.98	5.3
MCV-fl	1	303	302	92.05	85.1	5.3	7.28	2.65	7.62
MCH-Pg	1	303	302	87.09	92.38	6.95	4.64	5.96	2.98
MCHC-g/dl	1	303	302	93.05	93.38	3.64	3.31	3.31	3.31
Plt. x10³/µl	1	303	302	93.71	91.06	3.64	5.63	2.65	3.31
ReticCount%	2	303	251	95.22	86.85	3.19	7.17	1.59	5.98
PS Assessment	3	303	249	Satisfactory :92.74%, Borderline Sat. :3.30%, Unsatisfactory :3.96%					

***Comments:**

1). **Among Lab (EQA) : PS Diagnosis not reported, Results acceptable.**

2). **Within Lab (IQA) : RETIC result is unacceptable, may be due to random/human error.**

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ± 2 : Acceptable, Z score ± 2 to ± 3 :Warning Signal, Z score $> \pm 3$: Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between " 0 to ± 2 " are texted in green colour. Z score value between " ± 2 to ± 3 " are texted in orange colour. Z score value $> \pm 3$ are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value ($0.3*SDPA$). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value ($0.3*SDPA$).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Manoranjan Mahapatra (Prof. & Head)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----