



CALIBRATION CERTIFICATE

Certificate No.	MES/24-25/VMHSM/GLB/INC/D01	Date of Issue	: 08/05/2024
		Calibration Done On	: 06/05/2024
		Calibration Due On	: 05/05/2025

Customer Name	Vivekananda Memorial Hospital, Saragur, Mysore.
---------------	--

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Incubator	Kemi/KIS 2- F10	13215	-	General Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Thermometer	Metravi / DTM 900	200923435	TVCSPL23/09/313-01	15/09/2024
2	Digital Thermo Hygrometer	Equinox & EQ - 308 CTH	20191602	TVCSPL 23/06/372 - 01	13/06/2024

The instruments used for calibration are under valid calibration and are traceable to National Standards.
 The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

PERFORMANCE DATA

Set Reading	Display Reading	Measured Reading	Tolerance
30.0 °C	30.0 °C	30.1 °C	± 0.3 °C
50.0 °C	50.0 °C	50.1 °C	± 0.3 °C
70.0 °C	70.0 °C	70.1 °C	± 0.3 °C
90.0 °C	90.0 °C	90.2 °C	± 0.3 °C
110.0 °C	110.0 °C	110.2 °C	± 0.3 °C

Calibration Status	Pass	✓	Fail
--------------------	------	---	------

Calibrated By

Hari Krishnan (Service Engineer)

Approved By



Binoj Thomas (Quality Manager)

NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services.