

GSTIN: 29ATZPG5568K1ZL



Omkar Enterprises

ONE STOP SHOP FOR SPECIALITY RESEARCH
Lab & Medical Equipments Sales / Service & Calibration

AUTHORISED DEALERS FOR **REMI GROUP** REMI SALES & ENGINEERING LTD.

SPECIALIST IN BLOOD BANK EQUIPMENTS 24x7

Simply - 27

CALIBRATION CERTIFICATE

Certificate No: 56789A

Date :

VIVEKANAND MEMORIAL HOSPITAL

Date: 21/09/2023

SARGUR H D KOTE

MYSORE DIST ----- Karnataka

Description of The Item: **PIPETS**

Range 100- 1000 Micro ml

Make **FIXAPETTES** No F20274

Details Of Equipment Used For Calibration

Sl. No	Description	Certified By	Certificate No.	Validity
1	Brass Weight box 1 mg to 200+ 200 gms, 500g +1 kg +5 kg	RRSL(Weights & Measures) Bangalore Govt of India	RRSL/B/365/O/407/2023-24 RRSL/B/364/O/406/2023-24	07/09/2023
2	ELECTRONIC Weighing MACHINE	RRSL(Weights & Measures) Bangalore Govt of India	RRSL /B/440/O/498/2023-24	30/09/2023

Observation

Sl. No.	Set Weight (g / Micro ml)	Measured Weight (micro)	Difference in (micro)	Remarks
1	100	100.0	0.0	OK
2	500	500.01	0.0	OK
3	1000	1000.0	0.1	OK

Next Calibration: Due Date: 20/03/2024

- Remarks:
1. Calibration traceable to National Standards.
 2. Overall Deviation is within specified accuracy hence, instrument is OK.
 3. Results are related only to the item calibrated.
 4. Maximum tolerance permissible is +/- 2 deg cel. & +/- 100 RPM.

Calibrated By

VIJAY



Approved By

VILAS R G

13/1, 4th Main Road, Lakshmi Devamma Block, Ganganagar, Bengaluru - 560 032.
Email: omkar.enterprises02@gmail.com Cell : 94488 82229, 90087 79558, 74110 27736, 7676042255

Shot on OnePlus