



CALIBRATION CERTIFICATE

Certificate No.	MES/24-25/VMHSM/GLB/CEN/D01	Date of Issue	: 08/05/2024
		Calibration Done On	: 06/05/2024
		Calibration Due On	: 05/05/2025

Customer Name **Vivekananda Memorial Hospital, Saragur, Mysore.**

DETAILS OF DEVICE UNDER CALIBRATION

Equipment Name	Make / Model	Serial No.	Asset ID	Location
Centrifuge	Remi / R - 8C	ZJN33776	-	General Lab

DETAILS OF MASTER EQUIPMENT USED

Sl No.	Master Equipment	Make / Model	Serial No.	Reference ID	Cal. Due
1	Digital Tachometer	Metravi / NCTM-2000	210532216	TVCSPL23/09/313-02	15/09/2024
2	Digital Thermo Hygrometer	Equinox & EQ - 308 CTH	20191602	TVCSPL 23/06/372 - 01	13/06/2024

The instruments used for calibration are under valid calibration and are traceable to National Standards.

Environmental Conditions	Temperature	27°C	Relative Humidity	55 %
---------------------------------	-------------	------	-------------------	------

PERFORMANCE DATA

Set Reading	Display Reading	Measured Reading	Tolerance
Minimum	1000 rpm	1001 rpm	± 5 rpm
Medium	3000 rpm	3001 rpm	± 5 rpm
Maximum	5250 rpm	5252 rpm	± 5 rpm

Calibration Status	Pass	✓	Fail	
--------------------	------	---	------	--

Calibrated By

Hari Krishnan (Service Engineer)

Approved By

Binoj Thomas (Quality Manager)

NB: Results reported are valid at the time of and under the stated conditions of measurements. This report shall not be reproduced without written approval by Medical Engineering & Services.