

CALIBRATION CERTIFICATE

**INSTITUTION : ASTRA NEST
DIAGNOSTICS Bangalore**

ANALYZER: DIRUI H-100	CALIBRATION DATE : 21.05.2024
S.NO:2200100H0912	NEXT CALIBRATION :20.05.2025

This is to certify that the analyser has been calibrated and maintenance has been performed.

S.No	PARAMETERS	RANGE	OBTAINED VALUE
1	INPUT VOLTAGE	110-240 VAC	231 VAC
2	WAVELENGTH		
	525	4100-5100	4388
	572	4100-5100	4476
	610	4100-5100	4582
	660	4100-5100	4628
3	CAL STRIP	OK	OK

GENWORKS HEALTH PVT LTD

Gamma Block, 5th Floor, Sigma Tech Park, Whitefield Main Road, Varthur Hobli, Bangalore 560066, India
CIN :U24230KA2015PTC078753 | www.genworkshealth.com

SCALAR VALUES

PARAMETERS	660nm	610nm	572nm	525nm
UBG	630	465	320	210
BIL	950	880	785	
KET	3575	3300	3060	2655
BLD	5300	3850	2200	1050
PRO	650	570	480	350
NIT	835			
LEU	5280	4870	4165	3100
GLU	970	880	650	405
SG	3215	2380	1900	1360
pH	2700	2135	1855	1685

GENWORKS HEALTH PVT LTD

 Gamma Block, 5th Floor, Sigma Tech Park, Whitefield Main Road, Varthur Hobli, Bangalore 560066, India
 CIN :U24230KA2015PTC078753 | www.genworkshealth.com



A WIPRO GE HEALTHCARE INVESTED COMPANY

Report Sign Off:

Calibration Done By: Mr.Toufiq

Designation: Circle Manager-Customer Support

Date & Sign: 06.02.2024

For
M. N. Lakshman
(M. NO. CLAKANDAN)

Note: Supportive date to be attached along with this certificate.

GENWORKS HEALTH PVT LTD

Gamma Block, 5th Floor, Sigma Tech Park, Whitefield Main Road, Varthur Hobli, Bangalore 560066, India
CIN :U24230KA2015PTC078753 | www.genworkshealth.com