



TMSCC

TESTING MACHINE SERVICE AND CALIBRATION CENTRE
Precision is Our Destination.....

An ISO 9001 : 2015 Certified company
Calibration of Various Instruments & Testing Machines...



CC - 3125

Form No.: TMSCC/R/23

CALIBRATION CERTIFICATE

Page 1 of 1

CALIBRATION CERTIFICATE NO.:		2024/06/888				
ISSUE DATE:		18-Jun-2024				
ULR No.:		CC312524000012394F				
1.0 ISSUED TO:		M/s: Doctor Ghosh Super Speciality Clinic And Diagnostic Centre 1st Floor, 2/5/16, Arabinda Arena, Block B, Khardaha, Kolkata - 700118.				
1.1 Service Request Form No.:		SRF/2024/06/17/02				
1.2 Service Request Date:		17-Jun-2024				
1.3 Location:		At Site				
1.4 Description identification of item to be calibrated:						
i Name:		Refrigerator		ii Make:		
iii Model / Type No.:		RR19A241BGS/NL/2022		iv S/L No.:		
v I.D.No.:		DGLAB/REF/01		vi Job Code No.:		
vii Range:		Mentioned in the results		viii Resolution:		
ix Accuracy:		As Per DKD R5 - 1		x End User:		
1.5 Full / Partial Calibration:		Full Calibration				
1.6 Applicable specification of item to be calibrated: Accuracy / permissible limit:		Not Specified				
1.7 Date of receipt of item:		17-Jun-2024		1.7 Date of calibration:		
1.9 Calibration due on:		16-Jun-2025		1.9 Frequency of calibration once in:		
2.1 Environmental condition during calibration:		Temperature: 25.5°C		12 Months		
		Humidity: 55% RH				
2.2 Basis of calibration:		SOP/1/06				
2.3 Traceability: Standards used for calibration are traceable to National Standards through NABL accredited Laboratory						
Name of Instrument		SI. No. / ID no.	Certificate No.	Lab Certificate No.	Calibrated On	Due On
Digital Temperature Indicator with Sensor(RTD)		(Indicator SI. No. - 18K588073) (Sensor Id No. - TMSCC/R/D/01)	TSC/23-24/13010-4	CC-2231	09-11-2023	08-11-2024

CALIBRATION RESULTS

Sl. No.	DUC set in °C	Observed Reading at Ref. Std. in °C	Mean reading in °C	Error in °C	Uncertainty in ± °C	Acceptance Criteria
Normal Zone : 7 °C						
1	7	7.010	7.014	0.014	0.69	Pass
2		7.014				Pass
3		7.018				Pass
4		7.012				Pass
4		7.014				Pass
Deep Zone : (-)13 °C						
1	-13	-13.028	-13.029	-0.029	0.69	Pass
2		-13.033				Pass
3		-13.027				Pass
4		-13.029				Pass
5		-13.030				Pass

Maximum Permissible Error : ± 0.3% of rdg

Measurement Uncertainty at 95% Confidence Level where Coverage Factor k = 2

REMARKS : The DUC has been calibrated over its range. The readings observed are tabulated above. The reference standard is traceable to National standard.

DUC Device Under Calibration.

Physical Status of the DUC : OK

Calibrated By:

C. Ghosh
(Calibration Engineer)

Checked By:

C. Ghosh
(Technical Director)

Approved By:

P. K. Modak
(Sr. Calibration Engineer)

END OF CERTIFICATE