

Urine Routine Analysis

ALL METHOD REPORT

Cycle - 13/2024

Date: 31/08/2024

Lab Code: 3549

Round No -08

| Parameters | Units | No. of Participants | Consensus | | Consensus Result | Your Value | Remark |
|--------------------|-------|---------------------|-----------|-----|------------------|------------|-------------------|
| | | | Number | % | | | |
| Specific Gravity | - | 466 | 308 | 66% | 1.020 | *1.010 | Outside Consensus |
| pH | - | 466 | 315 | 68% | 7.0 | 7.0 | Within Consensus |
| Leukocytes | /μl | 449 | 303 | 68% | 2+ - 3+ | 3+ | Within Consensus |
| Nitrite | - | 442 | 384 | 87% | Positive | Positive | Within Consensus |
| Protein | mg/dl | 467 | 375 | 80% | 2+ - 3+ | *4+ | Outside Consensus |
| Glucose | mg/dl | 467 | 346 | 74% | 3+ - 4+ | 4+ | Within Consensus |
| Ketones | - | 460 | 297 | 65% | 2+ - 3+ | 2+ | Within Consensus |
| Urobilinogen | mg/dl | 459 | 337 | 73% | 2+ - 3+ | 3+ | Within Consensus |
| Bilirubin | mg/dl | 459 | 357 | 78% | 3+ | 3+ | Within Consensus |
| Blood (Hemoglobin) | /μl | 461 | 342 | 74% | 3+ - 4+ | 4+ | Within Consensus |

| Legends | ≥ 60% Consensus | (*) Outside Consensus | (@) No Consensus | (.) Not Reported | (#)Late Result Submission |
|---------|-----------------|-----------------------|------------------|------------------|---------------------------|
|---------|-----------------|-----------------------|------------------|------------------|---------------------------|

Chief Coordinator



Dr. Sanjay Mehrotra


Prepared By: NF

Programme Director



Dr. Bandana Mehrotra

Checked By:



Doc. No.: ASS / FR / 10 / R 01 / Dt.: 05.01.2024

Urine Routine Analysis

METHOD WISE REPORT PART-01

Cycle - 13/2024

Date: 31/08/2024

Lab Code: 3549

Round No -08

| Parameters | Units | Method Group | No. of Participants | Consensus | | Consensus Result | Your Value | Remark |
|--------------------|-------|----------------|---------------------|-----------|-----|------------------|------------|-------------------|
| | | | | Number | % | | | |
| Specific Gravity | - | Semi-Automated | 184 | 121 | 66% | 1.020 | *1.010 | Outside Consensus |
| pH | - | Semi-Automated | 184 | 129 | 70% | 7.0 | 7.0 | Within Consensus |
| Leukocytes | /µl | Semi-Automated | 184 | 144 | 78% | 2+ -3+ | 3+ | Within Consensus |
| Nitrite | - | Semi-Automated | 182 | 164 | 90% | Positive | Positive | Within Consensus |
| Protein | mg/dl | Semi-Automated | 184 | 154 | 84% | 2+ -3+ | *4+ | Outside Consensus |
| Glucose | mg/dl | Semi-Automated | 184 | 139 | 76% | 3+ - 4+ | 4+ | Within Consensus |
| Ketones | - | Semi-Automated | 184 | 121 | 66% | 1+ - 2+ | 2+ | Within Consensus |
| Urobilinogen | mg/dl | Semi-Automated | 184 | 139 | 76% | 2+ - 3+ | 3+ | Within Consensus |
| Bilirubin | mg/dl | Semi-Automated | 184 | 152 | 83% | 3+ | 3+ | Within Consensus |
| Blood (Hemoglobin) | /µl | Semi-Automated | 184 | 134 | 73% | 3+ - 4+ | 4+ | Within Consensus |

| | | | | | |
|---------|-----------------|-----------------------|------------------|------------------|---------------------------|
| Legends | ≥ 60% Consensus | (*) Outside Consensus | (@) No Consensus | (.) Not Reported | (#)Late Result Submission |
|---------|-----------------|-----------------------|------------------|------------------|---------------------------|

Chief Coordinator

Programme Director



Dr. Sanjay Mehrotra



Dr. Bandana Mehrotra

Prepared By: SM

Checked By: 