

Urine Routine Analysis

ALL METHOD REPORT

Cycle - 13/2024

Date: 31/07/2024

Lab Code: 3549

Round No -07

Parameters	Units	No. of Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	453	283	63%	1.020	*1.005	Outside Consensus
pH	-	452	307	68%	7.0	7.0	Within Consensus
Leukocytes	/μl	441	298	68%	2+ - 3+	3+	Within Consensus
Nitrite	-	434	353	81%	Positive	Positive	Within Consensus
Protein	mg/dl	452	350	77%	2+ - 3+	3+	Within Consensus
Glucose	mg/dl	453	315	70%	3+ - 4+	4+	Within Consensus
Ketones	-	447	271	61%	2+ - 3+	2+	Within Consensus
Urobilinogen	mg/dl	444	332	75%	2+ - 3+	3+	Within Consensus
Bilirubin	mg/dl	445	338	76%	3+	3+	Within Consensus
Blood (Hemoglobin)	/μl	448	315	70%	3+ - 4+	*5+	Outside Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	---------------------------

Chief Coordinator



Dr. Sanjay Mehrotra

Prepared By: NF

Checked By: 

Programme Director



Dr. Bandana Mehrotra

Doc. No.: ASS / FR / 10 / R 00 / Dt.: 05.01.2024