

Urine Routine Analysis

ALL METHOD REPORT

Cycle - 13/2024

Date: 29/06/2024

Lab Code: 3549

Round No -06

Parameters	Units	No.of. Participants	Consensus		Consensus Result	Your Value	Remark
			Number	%			
Specific Gravity	-	399	372	93%	1.010 - 1.020	1.020	Within Consensus
pH	-	399	358	90%	6.0	6.0	Within Consensus
Leukocytes	/μl	387	378	98%	Negative	Negative	Within Consensus
Nitrite	-	381	376	99%	Negative	Negative	Within Consensus
Protein	mg/dl	399	389	98%	Negative	Negative	Within Consensus
Glucose	mg/dl	399	393	99%	Negative	Negative	Within Consensus
Ketones	-	395	394	100%	Negative	Negative	Within Consensus
Urobilinogen	mg/dl	393	385	98%	Negative	Negative	Within Consensus
Bilirubin	mg/dl	393	376	96%	Negative	Negative	Within Consensus
Blood (Hemoglobin)	/μl	395	384	97%	Negative	Negative	Within Consensus

Legends	≥ 60% Consensus	(*) Outside Consensus	(@) No Consensus	(.) Not Reported	(#)Late Result Submission
---------	-----------------	-----------------------	------------------	------------------	---------------------------

Chief Coordinator



Dr. Sanjay Mehrotra

Prepared By: DP

Programme Director



Dr. Bandana Mehrotra

Checked By: 

Doc. No.: ASS / FR / 10A / R 00 / Dt.: 05.01.2024