

Calibration Certificate

Customer Name	ESIC Hospital
Address	Bibavewadi, Pune
Contact Person	Mr, Prafull Sagar
Contact No.	9029401791
Instrument	Hematology Analyzer
Model	BC 6200
Serial No.	TW-91000341
Instrument Status	Working
Calibration Done On	15.07.2024
Next Calibration Due On	14.07.2025

For Aspen Diagnostics Pvt. Ltd.

Area Service Manager



Prasad Jadhav

Calibrator Calibration


Date:-15.07.24

Instrument - 6 Part Haematology Analyzer	Model: BC-6200	Serial No.: TW-91000341
---	-------------------	----------------------------

Parameter	WBC	RBC	HGB	MCV	PLT
TARGET	9.33	4.53	13.0	92.0	249
1.	10.80	4.68	13.8	90.2	279
2.	10.81	4.70	13.8	90.3	288
3.	11.33	4.64	13.7	89.5	271
4.	11.06	4.66	13.8	90.1	272
5.	11.09	4.67	13.7	89.6	278
6.	10.95	4.67	13.8	89.5	262
7.	10.87	4.57	13.7	89.4	272
8.	11.24	4.63	13.8	89.8	276
9.	10.98	4.70	13.8	89.8	261
10	11.33	4.66	13.8	89.2	285
Mean	11.05	4.66	13.8	89.7	274
CV(%)	1.815	0.812	0.45	0.41	3.2
Old Cal.Factor (%)	82.63	96.00	94.00	100.00	100.00
New Cal.Factors (%)	84.46	97.23	94.44	102.52	90.76

For Aspen Diagnostics Pvt. Ltd.

Area Service Manager



Prasad Jadhav