

# Aspen Diagnostics Pvt. Ltd.

(Rapid Diagnostic Group Of Companies)




## Calibration Certificate

Customer Name	Payas Pathology Labrotory
Address	Dehugaon Pune, Maharashtra
Contact Person	Payas Pathology Labrotory
Contact No.	9146518783
Instrument	Urine Analyzer
Model	Uri Plus- 200
Serial No.	2300100H0345
Instrument Status	Working
Calibration Done On	12/03/2024
Next Calibration Due On	11/03/2025

For Aspen Diagnostics Pvt. Ltd.

Service Engineer



Prasad Jadhav

## Quality certificate of Urine Analyzer

Name	Payas Pathology Lab.
Hospital	Payas Pathology Lab.
Date	10.09.2024

Instrument identification				QC Passed	
Serial no.	2300100H0345	Date of delivery	12.03.24	Yes	Yes
Model	Uri plus200	Date of QC	10.09.24	yes	Yes

**InspectionLIST** (Relevant information should be filled in)

	YES/NO
1. The Voltage of Instrument should be 220 V $\pm$ 10%.	Yes
2. Ground Voltage should be less than 10 volts.	Yes
3. The instrument is tested Low and High Control.	Yes
4. Calibration passed.	Yes
5. Self testing passed.	Yes
6. Strip bed cleaned.	Yes
7. Fresh urine sample used.	Yes

Quality Control Result	Level	Target	Result	Remark
UBG	Positive Control		Positive Control	OK
BIL		3+	OK	
KET		2+	OK	
BLD		2+	OK	
PRO		2+	OK	
NIT		POS	OK	
LEU		3+	OK	
GLU		4+	OK	
SG		1.025	OK	
pH		6.0	OK	

Test results is coming ok	✓	Remarks	Done
---------------------------	---	---------	------

  
 Signature of Technical Person