



IN-HOUSE CALIBRATION CERTIFICATE

Name of the Equipment :- Centrifuge	
Equipment I.D. :- PPTCT/DWH/CENTRI/01	
Location :- Room No. 104, PPTCT District Women Hospital, Gorakhpur	Working Range:- 5000
Manufacture:- Oscar	
Calibrated on:- 01/07/2024	Next Calibration Due:- 30/06/2025

Method:- Traceability Certificate

MICRO/HDRL/Tach-01 SRF No. & Date: 200913A/N-105 Dt. 14/09/2023
Standard: Line ACS/DTM/01/211-0155V, G-230908-1-1 Dt. 08/09/2024

Test condition: Working

Room Temperature : 25 °C

RH: 65%

Instrument	Reading					Mean
	1	2	3	4	5	
Tachometer (BRDMC/Micro/HDRL/Tachometer/01/SRL)	3012	3021	3016	3020	3019	3017
Centrifuge 01 (PPTCT/DWH/CENTRI/01)	3000	3000	3000	3000	3000	3000
Acceptable variation: +/- 10% with standard Tachometer						

Remarks: Instrument has passed the acceptance criteria (Yes/No)..... YES

Jyoti Kumar
11/7/24
Checked by
Jyoti Kumar
Technical Manager
HIV Diagnostic & Research Laboratory
Department of Microbiology
B.R.D. Medical College Gorakhpur, (U.P.) 273013

Abhishek
11/7/24
Verified by
Lab Director/Quality Manager
HIV Diagnostic & Research Laboratory
Department of Microbiology
B.R.D. Medical College Gorakhpur, (U.P.) 273013