



IN-HOUSE CALIBRATION CERTIFICATE

Name of the Equipment :- Digital Thermometer	
Equipment I.D. :- PPTCT/DWH/THERMO/01	
Location :- Room No. 104, PPTCT District Women Hospital, Gorakhpur	Working Range:- -50°C
Manufacture:- Aptechdeals	
Calibrated on:- 01/07/2024	Next Calibration Due:- 30/06/2025

Method:- Traceability Certificate

MICRO/HDRL/Tach-01 SRF No. & Date: 220913A/N-103 Dt. 13/09/2023
Standard: Line ACS/TIS/01/20C106914, BCL30059991 Dt. 20/07/2024

Test condition: Working


Room Temperature : 25°C

RH: 65%

Instrument	Reading					
	1	2	3	4	5	Mean
Digital Thermometer SRL (BRD/Micro/HDRL/DegitalTerm./01/SRL)	36.3	37.1	37.1	36.9	36	36.68
Thermometer (HCTS) (PPTCT/DWH/DEGITALTERM./01)	36.9	37.1	36.5	36.2	36.6	36.66
Acceptable variation: +/- 10% with standard Tachometer						

Remarks: Instrument has passed the acceptance criteria (Yes/No)..... YES


Checked by
Jyoti Kumar
 Technical Manager
 HIV Diagnostic & Research Laboratory (HDRL)
 Department of Microbiology
 B.R.D. Medical College Gorakhpur, (U.P.) 273013


Verified by
Lab Director/Quality Manager
 HIV Diagnostic & Research Laboratory (HDRL)
 Department of Microbiology
 B.R.D. Medical College Gorakhpur, (U.P.) 273013