



MEASURE TECHNO LAB

2, B.T. ROAD (JAYANTI CINEMA COMPLEX), BARRACKPORE,
KOLKATA - 700120, W.B.



CC - 2545

Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974,
LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in

CALIBRATION CERTIFICATE OF MICRO PIPETTE

BRATION CERTIFICATE NO.: MTL / D / R09 / 06 - 23

ULR - CC254523000018173F

Page: 1 of 1

0 Service Request No.: MTL / 14F / 06 / 22 - 23

1 Issued to: M/s. Diagnostics,
190B, Rashbehari Avenue,
Kolkata - 700029.

2 Description & Identification of Item to be Calibrated:	a) Name:	Micro Pipette	b) Code No.:	DIA / BIO / MP - 03
	c) Sl. No.:	N.S.	d) Make:	Precisa
	e) Model / Type:	N.S.	f) Range:	20 µl to 200 µl
	g) Sensor:	N.A.	h) Resolution:	1 µl
	i) End User:	Biochemistry	j) Accuracy:	N.S.
	k) Calibration done at:	On Site / <input checked="" type="checkbox"/> In House		

3 Date of receipt of Item :	24-06-23	1.4 Physical Condition of DUC :	OK
5 Date of calibration :	26-06-23	1.6 Recommended date of next calibration :	26-06-24
7 Date of Issue :	28-06-23		

8 Environmental Conditions During Calibration:

Temperature:	20 °C ± 2 °C
Humidity:	30 % RH to 75 % RH
Pressure:	998.1 mbar

9 Method of Calibration: SOP / MASS / 02 (As Per ISO : 8655 - 6 : 2022)

10 Traceability :

- a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.
- b) The following standards / Equipment have been used.
 - Weights Cal. Certificate No. NC-210 (NSTAR, Ahmedabad) (Cal. Date: 16/05/22, Due Date: 15/05/25)
 - Digital Thermometer With Sensor Cal. Certificate No. MTL/TH/DTM/R01/09-22 (MTL, Barrackpore) (Cal. Date: 16/09/22, Due Date: 16/09/23)

11 Result :

Mechanical Calibration : (Volume)

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1.	Volume	20	20.4601	20.5419	0.5419	0.39
2.	20 µl to 200 µl	100	98.5783	98.9726	-1.0274	0.39
3.		200	200.3714	201.1729	1.1729	0.44

- Remarks: i) Cubical Expansion co-efficient of pipette material taken as $10^{-5} \mu\text{l} / ^\circ\text{k}$.
- ii) This result has an expanded uncertainty with a coverage factor $k=2$ at approximately 95% confidence level.
- iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.
- iv) Calibrations are carried out without any adjustment or repair.

DUC - Device Under Calibration N.S. - Not Specified N.A. - Not Applicable

Opinions and Interpretations	
Calibrated	Accepted / Valid for use
Limited Use	Rejected / Out of use

Calibrated by:
Calibration Engineer
Measure Techno Lab
Kolkata
Calibration Engineer

Checked / Approved by:
KOLKATA
Quality & Technical Manager
S. Pandey