



# CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004. INDIA

## DEPARTMENT OF CLINICAL BIOCHEMISTRY CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

### MONTHLY SUMMARY REPORT - AUGUST 2024

LAB NO: 7820

LAB NAME: **DISTRICT HOSPITAL REWA**

CONSTITUENT GROUP

**HbA1c**

Date of Result Entered: 10/08/2024

Date of Result Published: 02/09/2024

PT Item: Lyophilized human whole blood based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	HbA1c	HPLC-Ion Exchange	Biorad ( D10 / D100 / Variant Turbo )	705	5.57	4.15	0.23	5.5 %	-0.3	0.02

u\* - method of uncertainty

Z-Score	Interpretation
$ z  \leq 2.0$	Acceptable
$2.0 <  z  < 3.0$	Warning Signal
$ z  \geq 3.0$	Unacceptable (Action Signal)

#### LAB ADDRESS

DISTRICT HOSPITAL REWA,  
KUSHABHAU THAKRE DH REWA,  
Rewa,  
Madya Pradesh

#### Coordinator Contact Details:

E-Mail: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact No.: 0416-2283102

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential.  
CMC EQAS does not sub contract any components.  
\*\*\*\*\* End of Report \*\*\*\*\*

*Pamela Christudoss*

Dr. Pamela Christudoss,  
CMC EQAS Coordinator,  
Christian Medical College, Vellore.