



BIO-RAD LABORATORIES (INDIA) PRIVATE LIMITED

Site name:	DISTRIC HOSPITAL REWA
Address:	KUSHA BHAI THAKRE DISTRICT HOSPITAL REWA
Province-Postal code	486001
Contact Name	MR SHIVENDRA SINGH

Calibration Certificate

Date <small>(dd mmmm, yyyy)</small>	18-09-2024	Instrument Model	D-10 Hb Analyzer	Instrument Serial Number	DM0G522388
---	------------	-------------------------	------------------	---------------------------------	------------

We certify that the Bio-Rad instrument specified above and installed in your lab is calibrated and the performance was found satisfactory.

High Pressure Pump Flow Rate Reference: Service Manual: Section 7.7.3

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
Measuring cylinder hex base 10ML	2310100	06.06.24

Specified Range@ 1.5ml/Minute	Measured Value
1.425 – 1.575 ml/minute	1.55

Column Holder temperature Reference: Service Manual: Section 7.9

Measured with <small>(Device manufacturer and model)</small>	Device Serial Number	Device Calibration date <small>(dd mmmm, yyyy)</small>
FLUKE 179	26030078	06.06.24

Specified Range @ 35°C	Measured Value
33.4°C – 36.4°C	34.4

Detector log value Reference: Service Manual: Section 7.2 or Field Bulletin D-10-081

Measured with	Specified Range	Measured Value
Instrument Service Software	0.018 – 0.022 AU	0.021AU

The next calibration is due on	17-09-2025	<small>(dd mmmm, yyyy)</small>
--------------------------------	------------	--------------------------------

For Bio-Rad Laboratories India Pvt Limited

Authorized Signatory:


