



ASIAN TECHNOLOGY

(A Unit of Asiandot Technology Pvt. Ltd.)

(A HOUSE OF CALIBRATION)



CC-4012

CALIBRATION CERTIFICATE

Customer name And Address	M/S: Kushabhau Thakrey District Govt,Hospital Rewa Central Processing Lab Bichhiya Rewa Mp-486001	Service request No. & date	P/03/17/09/24&17/09/2024
		ULR no.	CC401224000001103F
		Cert. No.	AT/24000001103
		Date of Receipt of DUC	17/09/2024
		Date of calibration	17/09/2024
		Date of issue	18/09/2024
		Suggested due date	16/09/2025

Instrument Details			
Instrument name	Refrigerator/Freezer	S. No.	204680116360609717
Make	Whirpool	Location	Lab
Range / Size	-15 to 8 °C	Accuracy	-----
Least Count	0.1 °C	Visual Inspection	OK
I.D. No.	DH/REWA/CPL/REF/30	Calibration Performed at	Site

Detail of reference standards & Major equipments used	
Equipment Name	SPRT With Indicator
Make	Tempsans & Fluke
Model / SR.No.	PT-100X1 & 1586A / 05640
Certificate No.	CC350823000000098F
Calibration Validity	25/10/2024
Calibration by	S.B International

Environmental Condition	Temperature	(25±4)°C	Calibration Reference	ITS:90 (1990)
	Relative Humidity	(50±20) %	Work Instruction	AT/WI-10TH

Calibration Results			
Serial No.	Std. Reading in (°C)	Measured Reading in (°C)	Uncertainty At 95% C.L. (coverage factor k=2)
01.	-10.232	-10.0	±0.5 °C
02.	-5.324	-5.1	
03.	0.235	0.2	
04.	2.215	2.2	
05.	3.314	3.2	
06.	4.425	4.1	
07.	6.23 6	6.1	
08.	7.874	8.0	

Remarks:

<ul style="list-style-type: none"> ❖ (1) Standard equipment use for calibration are traceable to national/ international standards. ❖ (2)The reported expended 'uncertainty' of measurement is stated as the standard uncertainty of measurement multiplied by coverage factor k = 2 such that the coverage probability corresponds to approximately 95%.(3)The above results are valid at the time of and under the stated conditions measurement.(4)This certificate is refers only to the particular item submitted for calibration. (5) Next calibration due date given as requested by the customer.

Calibrated By
(Calibration-Engg./TM)
(DIPANRAJ)
Form No.QF-47



Approved By
(QM/TM)
(NEERAJ TYAGI)
Page No. 1 of 1

---End of Report---