



# MEASURE TECHNO LAB

2, B.T. ROAD (JAYANTI CINEMA COMPLEX), BARRACKPORE,  
KOLKATA - 700120, W.B.



CC - 2545

Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974,  
LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in

## CALIBRATION CERTIFICATE OF DIGITAL THERMO HYGROMETER

CALIBRATION CERTIFICATE NO.: MTL / HE / R04 / 09 - 24

ULR - CC254524000045112F

Page: 1 of 1

1.0 Service Request No.: MTL / 13A / 09 / 24 - 25

1.1 Issued to: M/s. Halder Enterprise,  
458 / 1A, Motilal Colony,  
P.O. - Rajbari, P.S. - Airport, Pin - 700081.

1.2 Description & Identification of item to be Calibrated:	a) Name:	Digital Thermo Hygrometer	b) Code No.:	HE / DTH - 03
	c) Sl. No.:	N.S.	d) Make:	N.S.
	e) Model / Type:	HTC - 2	f) Range:	(-) 50 °C to 70 °C & 20 % RH to 95 % RH
	g) Sensor:	N.A.	h) Resolution:	0.1 °C & 1 % RH
	i) End User:	Laboratory	j) Accuracy:	N.S.
	k) Calibration done at:	On Site / √ In House		

1.3 Date of receipt of item : 13-09-24

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 14-09-24

1.6 Recommended date of next calibration : 14-09-25

1.7 Date of Issue : 16-09-24

1.8 Environmental Conditions During Calibration: Temperature: 25 °C ± 4 °C  
Humidity: 30 % RH to 75 % RH

1.9 Method of Calibration: SOP / TH / HU - 01

### 2.0 Traceability :

- a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.
- b) The following standards / Equipment have been used.

i) Temperature & Humidity Meter Cal. Certificate No. RPL-CF-2024-THER-CSRF-0674-C-1192 (JRPM, Chennai) (Cal. Date: 15/07/24, Due Date: 14/07/25)

### 2.1 Result :

Thermal Calibration : (Specific Heat & Humidity.)

Sl. No.	Parameter/ Range	Measured Value on DUC °C	Corrected Std. Value °C	Error °C	Measurement Expanded Uncertainty ± °C & % RH
At 50 % RH					
1.	Temperature (-) 50 °C to 70 °C	10.3	10.59	-0.29	0.89
2.		29.8	30.17	-0.37	
3.		39.7	40.14	-0.44	
4.		48.5	49.13	-0.63	
At 25 °C					
	Humidity 20 % RH to 95 % RH	% RH	% RH	% RH	1.81
1.		25	25.85	-0.85	
2.		49	50.53	-1.53	
3.		78	80.98	-2.98	
4.		88	90.51	-2.51	

- Remarks: i) This result has an expanded uncertainty with a coverage factor k=2 at approximately 95% confidence level.  
 ii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.  
 iii) Error = DUC Reading - Standard Reading.  
 iv) Calibrations are carried out without any adjustment or repair.

DUC - Device Under Calibration      N.S. - Not Specified      N.A. - Not Applicable      Std. - Standard

Opinions and Interpretations	
Calibrated	√ Accepted / Valid for use
Limited Use	Rejected / Out of use

Calibration Engineer  
Measure Techno Lab  
Kolkata  
J. Dey  
Calibration Engineer

Checked / Approved by:  
  
Quality / Technical Manager  
S. Pandey