




CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - APRIL 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBAJY KNJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 21/04/2024

Date of Result Published: 07/05/2024

PI Item: Lympholized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Patients	AV	Participants		Your Value	Z Score	CV
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1640	385.53	10.48	40.34	328.37 mg/dL	1.08	1.99
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	656	64.58	11.79	7.62	64.97 mg/dL	0.05	0.52
3	CREATININE II	Jaffé End point II	Any Analyser (Automation / Semi Automation)	1233	6.68	11.71	0.78	7.05 mg/dL	0.5	0.04
4	T.B. LIRURIN II	Diazonium Salt / Colorimetric / Jendrassik II	Any Analyser (Automation / Semi Automation)	1436	2.6	16.3	0.42	2.32 mg/dL	-0.87	0.02
5	T-PROTEIN II	Bialkal - Colorimetric II	Any Analyser (Automation / Semi Automation)	1251	4.33	12.37	0.54	3.78 g/dL	-1.02	0.03
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1285	2.85	11.67	0.31	2.15 g/dL	-1.61	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1238	4.3	14.99	0.14	4.06 mg/dL	-0.88	0.04
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1453	80.3	11.84	10.89	81.91 mg/dL	-0.78	0.56
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point I	Any Analyser (Automation / Semi Automation)	1296	81.75	17.62	14.41	65.16 mg/dL	-1.15	0.8




CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME



MONTHLY SUMMARY REPORT - APRIL 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBAJIV KNUJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 21/04/2024

Date of Result Published: 07/05/2024

PT Item: Lysurilized human serum based

S. No	Analyte	Method / Principle	Analysed	No. of Participants	AV	Participants		Your Value	Z Score	L*
						CV	SDPA			

Z Score	Interpretation
z ≤ 2.0	Acceptable
2.0 < z < 3.0	Warning Signal
z ≥ 3.0	Unacceptable (Action Signal)

L* - method of uncertainty

LAB ADDRESS

GHSPL SAMBAJIV KNUJ HEALTH CARE LLP,
115, TAMBOLA, KANSHAWANIGAR, AN-08, GEP
HOTEL RAJVALI,
Pondicherry

Paula Christodou

Dr. Paula Christodou,
CMC EQAS Coordinator,
Christian Medical College, Vellore

Coordinator Contact Details:

E-MAIL: cmcaq@cmc.vcl.ac.in
Contact No: 0416 2282102

I verify accuracy and stability of the sample is passed
data in CMC EQAS reports in control list.
CMC EQAS does not audit methodology components.
***** End of Report *****