



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - MAY 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBHAJV KNJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 24/05/2024

Date of Result Published: 03/06/2024

PT Item: Lyophilized human serum based

S. No.	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1721	147.34	10.07	16.53	166.84 mg/dl	1.01	0.91
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	1141	57.61	12.24	6.49	59 mg/dL	0.47	0.43
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	954	2.78	12.5	0.47	2.34 mg/dl	0.41	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Any Analyser (Automation / Semi Automation)	1754	4.31	18.38	0.62	4.51 mg/dL	-0.8	0.06
5	T.PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	1526	5.21	09.6	0.61	5.33 g/dL	0.04	0.07
6	ALBUMIN II	BCG - Colorimetric II	Any Analyser (Automation / Semi Automation)	1321	4.2	11.15	0.35	3.1 g/dL	0.27	0.03
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	1609	7.04	14.79	1.04	5.45 mg/dL	-0.58	0.05
8	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	1757	116.63	12.35	13.73	116 mg/dL	-0.61	0.58
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	1294	113.13	15.87	20.54	117 mg/dL	0.08	0.74



CHRISTIAN MEDICAL COLLEGE

VELLORE, TAMIL NADU - 632004, INDIA

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME



PC-1024

MONTHLY SUMMARY REPORT - MAY 2024

LAB NO: 20400

LAB NAME: GHSPL SAMBHAJV KNJ HEALTH CARE LLP

CONSTITUENT GROUP

Chemistry II

Date of Result Entered: 24/05/2024

Date of Result Published: 03/06/2024

PT Item: Lyophilized human serum based

S. No	Analyte	Method / Principle	Analyser	No. of Participants	AV	Participants		Your Value	Z Score	u*
						CV	SDPA			

u* - method of uncertainty

Z-Score	Interpretation
$ z \leq 2.0$	Acceptable
$2.0 < z < 3.0$	Warning Signal
$ z \geq 3.0$	Unacceptable (Action Signal)

Pamela Christudoss

Dr. Pamela Christudoss,
CMC EQAS Coordinator,
Christian Medical College, Vellore.

LAB ADDRESS

GHSPL SAMBHAJV KNJ HEALTH CARE LLP,
BHATJANGLA, KRISHNANAGAR, NH-30, OPP
HOTEL HAVELI,
West Bengal

Coordinator Contact Details:

E-Mail: clinco@cmcvellore.ac.in
Contact No: 0416-2283102

Homogeneity and Stability of the sample is passed
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
**** End of Report ****